

CONFIRMATION

Notification: 0000025927_02_01_rev1
on changes according to EN 15267 regarding certificate 0000025927_02 dated 2 February 2015

Measuring system: AR500 with ER120 for NO₂, SO₂ and O₃

Manufacturer: Opsis AB
Skytteskogsvägen 16
244 02 Furulund
Sweden

German Federal Environmental Agency (UBA)

Announcement about the uniform practice in
monitoring emissions and ambient air.
25 February 2015
Federal Gazette BAnz AT 02 April 2015 B5


**IV. Notifications to the uniform practice for the continuous monitoring of
emission and ambient air:**


**13 Notification as regards Federal Environment Agency (UBA) notices of 23 February 2012
(Federal Gazette (BAnz.) p. 920, chapter IV number 2.1) and of 12 February 2013 (Federal
Gazette (BAnz) AT 05 March 2013 B10, chapter V 11th notification)**

The step motor for the automatic grid finding type RDM 543/100A, of manufacturer BERGER LAHR, in the AR500 measuring system with ER120 for NO₂, SO₂ and O₃, of the company Opsis AB, was discontinued and therefore replaced by the step motor for the automatic grid finding type RDM 545/100A of manufacturer BERGER LAHR.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 20 September 2014

TÜV Rheinland Energie und Umwelt GmbH
Cologne, 30. April 2015


i. A. Dipl.-Ing. Ruth Steinhagen-Pinnow


i. A. Dipl. Ing. Carsten Röllig

www.umwelt-tuv.de
teu@umwelt-tuv.de
Tel. +49 221 806-5200

TÜV Rheinland Energie und Umwelt GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body. This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.