

CONFIRMATION

Notification: 0000028729_02_02
on changes according to EN 15267 regarding certificate 0000028729_02 dated 21 January 2016

AMS designation: GM700-2 for HF

Manufacturer: SICK AG
Nimburger Str. 11
76276 Reute
Germany

German Federal Environment Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 13 July 2017
Publication: BAnz AT 31.07.2017 B12**

**II Notifications as regards the uniform practice for the continuous monitoring
of emission and ambient air:**

**19 Notification as regards Federal Environment Agency notices
of 17 July 2014 (BAnz AT 05.08.2014 B11, chapter I number 2.1) and
of 27 February 2016 (BAnz AT 14.03.2016 B7, chapter V, 29th notification)**

The GM700-2 measuring system for HF manufactured by SICK AG has been equipped with a new enclosure for the measuring head. This enclosure is 10mm wider and comes with a handle and a quick-release fastener. The servicing interface can now be accessed from the outside and insulation has been optimised.

The current software versions of the GM700-2 measuring system manufactured by SICK AG are:


9105060 YVB8 (measuring head)


9100821 YK10 (evaluation unit)

9091948 WJ24 (purge air)

Statement issued by TÜV Rheinland Energy GmbH dated 8 March 2017

TÜV Rheinland Energy GmbH
Cologne, 08 September 2017


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.