

# CONFIRMATION

Notification: 0000036946\_01\_01  
on changes according to EN 15267 regarding certificate 0000036946\_01 dated 18 July 2017

---

**AMS designation:** DUSTHUNTER T100 for Dust

**Manufacturer:** SICK Engineering GmbH  
Bergener Ring  
01458 Ottendorf Okrilla  
Germany

---

**German Federal Environment Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 13 July 2017  
Publication: BAnz AT 31.07.2017 B12**

**II. Notifications as regards the uniform practice for the continuous monitoring  
of emission and ambient air:**


**24 Notification as regards Federal Environment Agency (UBA) notices  
of 19 February 2009 (BAnz p. 899, chapter I number 1.5) and  
of 14 July 2016 (BAnz AT 01.08.2016 B11, chapter V, 10<sup>th</sup> notification)**


As an alternative light source to the XR-E LED the XM-L LED manufactured by the same manufacturer may be used for the DUSTHUNTER T100 measuring system monitoring dust manufactured by SICK Engineering GmbH. This change does not significantly affect the performance of the AMS.

Statement issued by TÜV Rheinland Energy GmbH dated 27 January 2017

---

TÜV Rheinland Energy GmbH  
Cologne, 08 September 2017

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.