

CONFIRMATION

Notification: 0000037051_01_01
on changes according to EN 15267 regarding certificate 0000037051_01 dated 18 July 2017

AMS designation: GM32 in-situ gas analyser cross-duct version for NO and SO₂

Manufacturer: SICK AG
Nimburger Str. 11
76276 Reute
Germany

German Federal Environment Agency (UBA)


**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 13 July 2017
Publication: BAnz AT 31.07.2017 B12**


- II. Notifications as regards the uniform practice for the continuous monitoring of emission and ambient air:**
- 17 Notification as regards Federal Environment Agency (UBA) notices of 3 August 2009 (BAnz p. 2929, chapter I number 3.2) and dated 22 February 2017 (BAnz AT 15.03.2017 B6, chapter V 27th notification)**

Deuterium lamp D2Plus-DS Type XD6201-01 (transmittance version) may now also be used for the GM 32 in-situ gas analyser in its cross-duct version measuring NO and SO₂ manufactured by SICK AG.

Statement issued by TÜV Rheinland Energy GmbH dated 8 March 2017

TÜV Rheinland Energy GmbH
Cologne, 08 September 2017


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.