

# CONFIRMATION

Notification: 0000038499\_00\_01\_rev1  
on changes according to EN 15267 regarding certificate 0000038499 dated 22 March 2013

---

**Measuring system:** FLOWSIC 100

**Manufacturer:** SICK Engineering GmbH  
Bergener Ring 27  
01458 Ottendorf-Okrilla  
Germany

---

## German Federal Environmental Agency (UBA)

**Announcement about the uniform practice in  
monitoring emissions and ambient air.**

**03 July 2013**

**Federal Gazette BAnz AT 23.07.2013 B4**

**IV. Notifications to the uniform practice for the continuous monitoring  
of emission and ambient air:**

**3 Correction of Federal Environmental Agency notice of 12 February 2013  
Federal Gazette (BAnz AT 05.03.2013 B10, chapter II, no. 2.2)**

The current MCU Firmware version of the Flowsic 100 AMS for velocity as manufactured by SICK Engineering GmbH should read as follows:

MCU Firmware: 01.08.00 (instead of 01.08.01)

---

Statement of TÜV Rheinland Energie und Umwelt GmbH of 28 May 2013

---

continuation next page

<a href="http://www.umwelt-tuv.de">www.umwelt-tuv.de</a> teu@umwelt-tuv.de Tel. +49 221 806-5200	TÜV Rheinland Energie und Umwelt GmbH Am Grauen Stein 51105 Cologne
--	---

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

**V. Notifications to the uniform practice for the continuous monitoring of emission and ambient air:**

**13 Notification as regards Federal Environmental Agency notices regarding performance tested AMS manufactured by SICK Engineering GmbH and by SICK AG with regard to notification 14 of this confirmation**

Sequential no. 7 The current software version of the SOPAS ET platform for AMS control is: SOPAS ET 2.38.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 25 March 2013

---

**14 Notification as regards Federal Environment Agency notices of 14 February 2008 (Federal Gazette (BAnz.) p. 901, chapter II, no. 1.1) and 12 February 2013 (Federal Gazette (BAnz) AT 05.03.2013 B10, chapter II, no. 2.2)**

**FLAWSIC 100 / SICK Engineering GmbH**

The current software versions for the Flawsic 100 AMS for exhaust gas velocity as manufactured by SICK Engineering GmbH are:


MCU Firmware:	01.08.00
Software Sensor:	- Typ Pr, PR AC und S: 21.04.16
	- Typ M, M-AC, H, H-AC, PM, PH, PH-S: 1.04.16

The software platform SOPAS ET in a certified version is required for full control of the measuring system.


Statement of TÜV Rheinland Energie und Umwelt GmbH of 26 March 2013

---

TÜV Rheinland Energie und Umwelt GmbH  
Cologne, 06 August 2013



i. A. Dipl.-Ing. Guido Baum



i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.de](http://www.umwelt-tuv.de)  
teu@umwelt-tuv.de  
Tel. +49 221 806-5200

TÜV Rheinland Energie und Umwelt GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body. This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.