

CONFIRMATION

Notification: 00000039320_00_01_rev1
on changes according to EN 15267 regarding certificate 00000039320 dated 20 August 2013

Measuring system: Model 49i for O₃

Manufacturer: Thermo Fisher Scientific
24 Forge Parkway
Franklin, MA 02038
USA

German Federal Environmental Agency (UBA)

Announcement about the uniform practice in
monitoring emissions and ambient air.

27 February 2014

Federal Gazette BAnz AT 01.04.2014 B12

**VI. Notifications to the uniform practice for the continuous monitoring of
emission and ambient air:**


**35 Notification of the announcements of the Federal Environment Agency
of 21 February 2006 (BAnz. p. 2653, Chapter IV Number 3.2) and
of 3 July 2013 (BAnz AT 23.07.2013 B4, Chapter V 22nd notification)**

The Arcturus Bd. 101491-xx processor board for the Model 49i measuring system for O₃ by Thermo Fisher Scientific has been withdrawn and replaced by the new Arcturus Bd. 110570-xx processor board.


The current software version of the measuring system is: V 02.00.00 (113421-00)

Statement of TÜV Rheinland Energie und Umwelt GmbH of 1 October 2013

TÜV Rheinland Energie und Umwelt GmbH
Cologne, 17 April 2014



i. A. Dipl.-Ing. Ruth Steinhausen



i. A. Dipl.-Ing. Karsten Pletscher

www.umwelt-tuv.de
teu@umwelt-tuv.de
Tel. +49 221 806-5200

TÜV Rheinland Energie und Umwelt GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.