

# CONFIRMATION

Notification: 0000039320\_00\_04  
on changes according to EN 15267 regarding certificate 0000039320\_00 dated 20 August 2013

---

**Measuring system:** Modell 49i for O<sub>3</sub>

**Manufacturer:** Thermo Fisher Scientific  
27, Forge Parkway  
Franklin, MA 02038  
USA

---

## German Federal Environmental Agency (UBA)

**Announcement about the uniform practice in  
monitoring emissions and ambient air.  
dated 14 March 2016  
Federal Gazette: BAnz AT 14.03.2016 B7**

**V. Notifications to the uniform practice for the continuous monitoring of  
emission and ambient air:**

- 15 **Notification as regards Federal Environment Agency (UBA) notices of  
21 February 2006 (BAnz. p. 2653, chapter IV number 3.2) and of  
25 February 2015 (BAnz AT 02.04.2015 B5, chapter IV notification 19)**

The current software version of the measuring equipment Model 49i for O<sub>3</sub>  
of Thermo Fisher Scientific is:


V 02.02.06


The EBM Pabst 8314 HL fan may be used instead of the one which is currently being  
used.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 22 October 2015

---

TÜV Rheinland Energy GmbH  
Cologne, 26 April 2016

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl. Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.