

CONFIRMATION

Notification: 0000040214_00_03
on changes according to EN 15267 regarding certificate 0000040214_00 dated 29 April 2014

Measuring system: Model 5014i Beta with PM₁₀-pre-seperator for particulate matter PM₁₀

Manufacturer: Thermo Fisher Scientific
27, Forge Parkway
Franklin, MA 02038
USA

German Federal Environmental Agency (UBA)

**Announcement about the uniform practice in
monitoring emissions and ambient air
dated 22 February 2017
Federal Gazette: BAnz AT 15.03.2017 B6**


V Notifications to the uniform practice for the continuous monitoring of emission and ambient air:


- 1 Notification as regards Federal Environment Agency notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 7.2) and
of 25 February 2015 (BAnz AT 02.04.2015 B5 chapter IV 24th notification)

The 5014i measuring system with PM₁₀ pre-separator for PM₁₀ suspended particulate
matter manufactured by Thermo Fisher Scientific may also be operated with a
GAST 87R647-PDS-HV-913 vacuum pump.

Statement issued by TÜV Rheinland Energie und Umwelt GmbH
dated 22 October 2015

TÜV Rheinland Energy GmbH
Cologne, 28 March 2017


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl. Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.