

# CONFIRMATION

Notification: 2704952ts\_01\_02  
on changes according to EN 15267 regarding certificate 2704952ts\_01 dated 09 June 2017

---

**AMS designation:** EL3000-Limas23 for NO, NO<sub>2</sub>, SO<sub>2</sub> and O<sub>2</sub>

**Manufacturer:** ABB Automation GmbH  
Stierstädter Str. 5  
60488 Frankfurt/Main  
Germany

---

**German Federal Environment Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 13 July 2017  
Publication: BAnz AT 31.07.2017 B12**

**II. Notifications as regards the uniform practice for the continuous monitoring  
of emission and ambient air:**

**5 Notification as regards Federal Environment Agency notices  
of 18 February 2016 (BAnz AT 14.03.2016 B7, chapter I number 4.2) and  
of 22 February 2017 (BAnz AT 15.03.2017 B6, chapter IV 6<sup>th</sup> correction and  
chapter V 30<sup>th</sup> notification)**

The current software version of the EL3000-Limas23 measuring NO, NO<sub>2</sub>, SO<sub>2</sub> and O<sub>2</sub>  
manufactured by ABB Automation GmbH is:


Limas23 (AMC-Board): 3.8.0


The Inpotron Type PSU-0261-12-14 power supply may be used as a 24V power supply.

Statement issued by TÜV Rheinland Energy GmbH dated 7 March 2017

---

TÜV Rheinland Energy GmbH  
Cologne, 08 September 2017

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.