

## CONFIRMATION

Notification: 0000053802\_02\_04 on changes according to EN 15267 regarding certificate 0000053802\_02 dated 02 March 2022

AMS designation:

ACF5000 for CO, NO, NO<sub>2</sub>, N<sub>2</sub>O, SO<sub>2</sub>, HCI, HF, NH<sub>3</sub>, CH<sub>4</sub>, CH<sub>2</sub>O, TOC, O<sub>2</sub>,

H<sub>2</sub>O and CO<sub>2</sub>

Manufacturer:

ABB AG

Stierstädter Str. 5 60488 Frankfurt/Main

Germany

## German Federal Environmental Agency (UBA)

Announcement regarding the uniform practice in monitoring emissions and ambient air dated 05 July 2023
Publication: BAnz AT 02.08.2023 B7

- III. Notifications as regards the uniform practice in monitoring emissions and ambient air:
- Notification as regards Federal Environment Agency (UBA) notices of 13 July 2017 (BAnz AT 31.07.2017 B12, chapter I number 2.1) and of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 1)

The current software versions for the ACF5000 measuring system for the components  $O_2$ , CO, NO,  $NO_2$ ,  $N_2O$ ,  $SO_2$ , HCI, HF,  $NH_3$ ,  $H_2O$ ,  $CO_2$ ,  $H_2CO$ ,  $CH_4$  and TOC from the company ABB AG are:

AMC board: 3.9.8 Syscon: 5.2.50

spectrometer: FTE 1.52

Statement issued by TÜV Rheinland Energy GmbH dated 31 March 2023

TÜV Rheinland Energy GmbH Cologne, 02 August 2023

i. V. Jol.-Ing. Guido Baum

i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu

Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH

Am Grauen Stein 51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.