

CONFIRMATION

Notification: 0000040204_02_03
on changes according to EN 15267 regarding certificate 0000040204_02 dated 01 July 2020

AMS designation: Serinus 40 for NO, NO₂ and NO_x

Manufacturer: ACOEM Australasia (Ecotech Pty Ltd)
1492 Ferntree Gully Road,
Knoxfield, VIC, 3180
Australia

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 21 February 2023
Publication: BAnz AT 20.03.2023 B6**

**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**59 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 4.1) and
of 9 March 2022 (BAnz AT 11.04.2022 B10, chapter VI notification 5)**

The current software version of the Serinus 40 measuring system for NO, NO₂ and NO_x from the company ACOEM Australasia for devices with the microprocessor board (C010001) remains: V 2.35.0001

The current software version of the Serinus 40 measuring system for NO, NO₂ and NO_x from the company ACOEM Australasia for devices with the microprocessor board (C010014) is: V 4.18.0000.


Furthermore, the following software versions are approved for this device version:
V 4.14.0000, V 4.15.0000, V 4.16.0000, V 4.17.0000


The software version number appears in the display of the measuring system in the format 2.XX or 3.XX or 4.XX.

Alternatively, the measuring system can also be operated with a converter filling made of molybdenum (Art. No.: C050019).

Statement issued by TÜV Rheinland Energy GmbH dated 05 September 2022

TÜV Rheinland Energy GmbH
Cologne, 25 April 2023


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.