

CONFIRMATION

Notification: 0000040205_02_03
on changes according to EN 15267 regarding certificate 0000040205_02 dated 01 July 2020

AMS designation: Serinus 50 for SO₂

Manufacturer: ACOEM Australasia (Ecotech Pty Ltd)
1492 Ferntree Gully Road,
Knoxfield, VIC, 3180
Australia

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 21 February 2023
Publication: BAnz AT 20.03.2023 B6**

**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**60 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 3.1) and
of 9 March 2022 (BAnz AT 11.04.2022 B10, chapter VI notification 6)**

The current software version of the Serinus 50 measuring system for SO₂ from the company ACOEM Australasia for devices with the microprocessor board (C010001) remains: V 2.35.0001


The current software version of the Serinus 50 measuring system for SO₂ from the company ACOEM Australasia for devices with the microprocessor board (C010014) is: V 4.18.0000.


Furthermore, the following software versions are approved for this device version:
V 4.14.0000, V 4.15.0000, V 4.16.0000, V 4.17.0000

The software version number appears in the display of the measuring system in the format 2.XX or 3.XX or 4.XX.

Statement issued by TÜV Rheinland Energy GmbH dated 05 September 2022

TÜV Rheinland Energy GmbH
Cologne, 25 April 2023


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.