

CONFIRMATION

Notification: 0000040206_03_01 on changes according to EN 15267 regarding certificate 0000040206_03 dated 02 June 2021

AMS designation: Spirant BAM 1000 mit PM₁₀-Vorabscheider for PM₁₀

Manufacturer: ACOEM Australasia (Ecotech Pty Ltd)

1492 Ferntree Gully Road, Knoxfield, VIC, 3180

Australia

German Federal Environmental Agency (UBA)

Announcement regarding the uniform practice in monitoring emissions and ambient air dated 21 February 2023
Publication: BAnz AT 20.03.2023 B6

- IV. Notifications as regards the uniform practice in monitoring emissions and ambient air:
- 61 Notification as regards Federal Environment Agency (UBA) notices of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 7.1) and of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 7)

The company name of Ecotech Pty. Ltd. has changed to ACOEM Australasia.

The Spirant BAM 1000 measuring system with PM_{10} pre-separator for suspended particulate matter PM_{10} from the company ACOEM Australasia will have a touch-sensitive screen in the future and a USB connection will now be installed in the front panel. On the rear panel there is now an electrical connection for a new generation of digital frahling sensors. In this context, the previous temperature sensor (BX-592) will be replaced with the digital model BX-598, or the combined pressure and temperature sensor BX-596 will be replaced with the digital model BX-597A, which also measures humidity.

The current software version of the measuring system is: 81237-05 V.1.1.0 or R9.4.0 (version with touch-sensitive screen).

Statement issued by TÜV Rheinland Energy GmbH dated 17 September 2022

TÜV Rheinland Energy GmbH Cologne, 25 April 2023

i. V. Mol.-Ing. Guido Baum

i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu

TRE@umwelt-tuv.eu Tel. +49 221 806-5200 TÜV Rheinland Energy GmbH

Am Grauen Stein 51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.