

CONFIRMATION

Notification: 0000035018_02_01
on changes according to EN 15267 regarding certificate 0000035018_02 dated 20 July 2022

AMS designation: Oxatex 3107 C67 for O₂

Manufacturer: Fives Pillard
Les Baronnies - BâtA / ZAC de Chateau Gombert, 3 rue Marc Donadille
13013 Marseille
France

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 28 June 2022
Publication: BAnz AT 28.07.2022 B4**


**III. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**


**59 Notification as regards Federal Environment Agency (UBA) notices
of 6 July 2012 (BAnz AT 20.07.2012 B11, chapter II number 1.1) and
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 29)**

The Oxatex 3107 C67 oxygen measuring device from Fives Pillard is also suitable for
use on plants complying with the 44th BImSchV.

Statement issued by TÜV Rheinland Energy GmbH dated 12 May 2022

TÜV Rheinland Energy GmbH
Cologne, 01 August 2022


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.