Umwelt 🎧 Bundesamt



# CERTIFICATE

of Product Conformity (QAL1)

Certificate No: 0000035018\_02

Certified AMS:	Oxatex 3107 C67 for O <sub>2</sub>
Manufacturer:	Fives Pillard Les Baronnies - BâtA / ZAC de Chateau Gombert, 3 rue Marc Donadille 13013 Marseille France
Test Institute:	TÜV Rheinland Energy GmbH

#### This is to certify that the AMS has been tested and found to comply with the standards EN 15267-1 (2009), EN 15267-2 (2009), EN 15267-3 (2007) and EN 14181 (2015).

Certification is awarded in respect of the conditions stated in this certificate (this certificate contains 7 pages). The present certificate replaces certificate 0000035018\_01 dated 18 July 2017.



Publication in the German Federal Gazette (BAnz) of 20 July 2012

German Environment Agency Dessau, 20 July 2022

how h

Dr. Marcel Langner Head of Section II 4.1

www.umwelt-tuv.eu tre@umwelt-tuv.eu Tel. + 49 221 806-5200 Suitability Tested EN 15267 QAL1 Certified Regular Surveillance

www.tuv.com ID 0000035018

This certificate will expire on: 19 July 2027

TÜV Rheinland Energy GmbH Cologne, 19 July 2022

Do Port hor

ppa. Dr. Peter Wilbring

TÜV Rheinland Energy GmbH Am Grauen Stein 51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

qal1.de

info@qal.de

page 1 of 7





Test report: Initial certification: Expiry date: Certificate:

Publication:

936/21212141/A dated 20 March 2012 20 August 2012 19 July 2027 Renewal (of previous certificate 0000035018\_01 of 18. Juli 2017 valid until 19 July 2022) BAnz AT 20.07.2012 B11, Chap. II No. 1.1

#### Approved application

The tested AMS is suitable for use at plants according to Directive 2010/75/EU, chapter III (13th BImSchV:2009), chapter IV (17th BImSchV:2009), 30th BImSchV:2009, Directive 2015/2193/EC (44th BImSchV:2021), TA Luft:2002 and at plants according to the 27th BImSchV:1997. The measured ranges have been selected so as to ensure as broad a field of application as possible.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a 3 month field test at a waste incineration plant.

The AMS is approved for an ambient temperature range of -20° to +50°C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the emission limit values and oxygen concentration relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the installation at which it will be installed.

#### Note:

The legal regulations mentioned do not correspond to the current state of legislation in every case. Each user should, if necessary, in consultation with the competent authority, ensure that this AMS meets the legal requirements for the intended use. In addition, it cannot be ruled out that legal regulations governing the use of a measuring device for emission monitoring may change during the lifetime of the certificate.

#### Basis of the certification

This certification is based on:

- Test report 936/21212141/A dated 20 March 2012 of TÜV Rheinland Energie und Umwelt GmbH
- Suitability announced by the German Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

# Umwelt 🎧 Bundesamt

#### Certificate: 0000035018\_02 / 20 July 2022



Publication in the German Federal Gazette: BAnz AT 20.07.2012 B11, chapter II No. 1.1, Announcement by UBA dated 06 July 2012:

# AMS designation:

Oxatex 3107 C67 for O2

#### Manufacturer:

Fives Pillard, Marseille, France

## Field of application:

For plants requiring official approval and for plants according to the 27th BImSchV

## Measuring ranges during the performance test:

Component	Certification range	Unit
O <sub>2</sub>	0 - 20.9	Vol%

## Software version:

V 3.1

## **Restrictions:**

None

#### Notes:

- 1. The maintenance interval is four weeks.
- 2. For peripheral parameters characterised by a high moisture content in combination with a high content of dust or a high dust content with a tendency of sticking, the filter needs to be checked more frequently.

#### **Test report:**

TÜV Rheinland Energie und Umwelt GmbH, Cologne Report No.: 936/21212141/A dated 20 March 2012





Publication in the German Federal Gazette: BAnz AT 02.04.2015 B5, chapter IV notification 32, Announcement by UBA dated 25 February 2015:

## 32 Notification as regards Federal Environment Agency (UBA) notice of 06 July 2012 (BAnz AT 20.07.2012 B11, chapter II number 1.1)

The current software version for the OXATEX 3107 C67 measuring system for oxygen, manufactured by Fives Pillard, is: V 3.3 (display) V 3.4 (CPU)

Statement of TÜV Rheinland Energie und Umwelt GmbH of 30 September 2014

Publication in the German Federal Gazette: BAnz AT 03.05.2021 B9, chapter III notification 29, Announcement by UBA dated 31 March 2021:

29 Notification as regards Federal Environment Agency (UBA) notices of 6 July 2012 (BAnz AT 20.07.2012 B11, chapter II number 1.1) and of 25 February 2015 (BAnz AT 02.04.2015 B11, chapter IV notification 32)

Fives Pillard, manufacturer of the Oxatex 3107 C67 oxygen measuring system, has relocated its headquarters. The new company address is: Fives Pillard Les Baronnies - Batiment A ZAC du Chateau Gombert 3 rue Marc Donadille 13013 Marseille France

Statement issued by TÜV Rheinland Energy GmbH dated 08 August 2020

#### **Certified product**

This certificate applies to automated measurement systems conforming to the following description:

The Oxatex 3107 C67 measuring system measures  $O_2$  in the measuring range of 0 - 20.9 Vol.-% using a zirconium dioxide probe.

The measuring instrument is comprised of a sensor with a heater, a flue gas inlet tube, a dissipater and the rear housing comprising electronics such as a display and an infra-red remote control. Additional optional test gas bottles with a suitable flow regulator are available.

info@gal.de





Test gas is brought to the sensor due to natural circulation of at least 0.5 m/s which is led into the inlet tube by a deflector situated at its tip.

The flue gas inlet tube is divided into two by the extension of the deflector, which forms an angle of 135° with the gas flow direction. This way, flue gas enters the inlet tube on the side facing the flue gas and leaves the tube on the opposite side.

The sensor is situated near the rear housing which is connected to the probe. The sensor comprises an electrode in contact with the gas to be measured and an electrode in contact with the reference gas. The test gas and the reference gas are separated by a layer of zirco-nium dioxide.

At high temperatures, zirconium dioxide becomes a conductor of oxygen ions.

The difference in oxygen content of the gasses on each side of the zirconium dioxide generates a voltage between the electrodes. According to Nernst's law, this voltage is proportional to the oxygen content of the gas being measured. The rear housing on the OXATEX 3107 C67, which is connected to the sensor, has the electronics required for sensor signal processing.

Energy supply is ensured via a ring lock circular connector at the front of the electronics housing. This is also where analogue and digital signal outputs are situated.

A dissipater situated between the flue gas inlet tube and the rear housing. This avoids an over-heating of the electronics housing.

The measuring system is configured using an infra-red remote control. For this purpose, a display board is integrated into the electronics housing.

The current oxygen concentration as Vol.% is displayed during the operation of the measuring system. Furthermore, signal lamps indicate if certain limit values are exceeded.

#### **General notes**

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document and the certification mark remains property of TÜV Rheinland Energy GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: **<u>gal1.de</u>**.





#### History of documents

Certification of Oxatex 3107 C67 is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

## Initial certification according to EN 15267

Certificate No. 0000035018\_00: 20 August 2012 Expiry date of the certificate: 19 July 2017 Test report 936/21212141/A dated 20 March 2012 TÜV Rheinland Energie und Umwelt GmbH Publication BAnz AT 20.07.2012 B11, chapter II number 1.1 UBA announcement dated 6 July 2012

#### Notifications

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 30 September 2014 Publication BAnz AT 02.04.2015 B5, chapter IV notification 32 UBA announcement dated 25 February 2015 (Software changes)

#### **Renewal of certificate**

Certificate No. 0000035018_01:	18 July 2017
Expiry date of the certificate:	19 July 2022

#### Notifications

Statement issued by TÜV Rheinland Energy GmbH dated 8 August 2020 Publication BAnz AT 03.05.2021 B9, chapter III notification 29 UBA announcement dated 31 March 2021 (Adress change)

# **Renewal of certificate**

Certificate No. 0000035018_02:	20 July 2022
Expiry date of the certificate:	19 July 2027

# Umwelt 🎧 Bundesamt

Certificate: 0000035018\_02 / 20 July 2022



## Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system				
Manufacturer	Fives Pillard			
Name of measuring system	Oxatex 3107 C67			
	09020117 / 09020118 / 10050138 / 10	050142/		
Serial number of the candidates	11090196 / 11090200			
Measuring principle	zirconium oxide			
Test report	936/21212141/A			
	TÜV Rheinland			
Test laboratory	2012-03-20			
Date of report	2012-03-20			
Measured component	O <sub>2</sub>			
Certification range	0 - 20.9 Vol%			
Evaluation of the cross sensitivity (CS)				
(system with largest CS)				
Sum of positive CS at zero point	0.00 Vol%			
Sum of negative CS at zero point	0.00 Vol%			
Sum of postive CS at reference point	0.17 Vol%			
Sum of negative CS at reference point	0.17 Vol%			
Maximum sum of cross sensitivities	0.17 Vol%			
Uncertainty of cross sensitivity	0.100 Vol%			
Calculation of the combined standard uncertainty		And the second second		
Tested parameter	U	U <sup>2</sup>		
Standard deviation from paired measurements under field conditions *	u <sub>D</sub> 0.088 Vol%	0.008 (Vol%) <sup>2</sup>		
Lack of fit	u <sub>lof</sub> 0.058 Vol%	0.003 (Vol%) <sup>2</sup>		
Zero drift from field test	u <sub>d.z</sub> 0.100 Vol%	0.010 (Vol%) <sup>2</sup>		
Span drift from field test Influence of ambient temperature at span	u <sub>d.s</sub> 0.120 Vol% u <sub>t</sub> 0.087 Vol%	0.014 (Vol%) <sup>2</sup>		
Influence of supply voltage		0.008 (Vol%) <sup>2</sup>		
Cross sensitivity (interference)		0.000 (Vol%) <sup>2</sup> 0.010 (Vol%) <sup>2</sup>		
Influence of sample pressure		0.010 (Vol%) <sup>2</sup>		
Uncertainty of reference material at 70% of certification range		0.029 (Vol%) <sup>2</sup>		
* The larger value is used :	u <sub>rm</sub> 0.169 Vol%	0.020 (10170)		
"Repeatability standard deviation at span" or				
"Standard deviation from paired measurements under field conditions"				
	$\left[\sum_{i=1}^{n} \left( \sum_{i=1}^{n} \right)^{2}\right]$			
Combined standard uncertainty (u <sub>C</sub> )	$u_{c} = \sqrt{\sum (u_{max, j})^{2}}$	0.30 Vol%		
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	0.60 Vol%		
Deletive total evenended uncertainty				
Relative total expanded uncertainty Requirement of 2000/76/EC and 2001/80/EC	U in % of the range 20.9 Vol%	2.9 10.0 **		
· ·	U in % of the range 20.9 Vol%	7.5		
Requirement of EN 15267-3	U in % of the range 20.9 Vol%	7.5		

\*\* For this component no requirements in the EC-directives 2001/80/EG und 2000/76/EG are given. A value of 10 % was used for this.