

CONFIRMATION

Notification: 0000025926_05_02
on changes according to EN 15267 regarding certificate 0000025926_05 dated 12 February 2020

AMS designation: MCS 100 FT for CO, NO, NO₂, N₂O, SO₂, HCl, HF, NH₃, CH₄, TOC, O₂, H₂O and CO₂

Manufacturer: SICK AG
Gisela Sick Straße 1
79276 Reute
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 21 February 2023
Publication: BAnz AT 20.03.2023 B6**


IV. Notifications as regards the uniform practice in monitoring emissions and ambient air:


38 Notification as regards Federal Environment Agency (UBA) notices of 10 January 2011 (BAnz. p. 294, chapter I number 3.1) and of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 53)

The MCS100FT measuring device for O₂, CO, SO₂, NO, NO₂, HCl, HF, CH₄, CO₂, H₂O, N₂O, NH₃ and TOC from SICK AG can also be operated with a revised gas sampling filter (SFU). This is mounted on the revised housing with protection class IP66.

Statement issued by TÜV Rheinland Energy GmbH dated 16 September 2022

TÜV Rheinland Energy GmbH
Cologne, 25 April 2023


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.