



## CERTIFICATE

of Product Conformity (QAL1)

Certificate No: 0000036946\_02

**Certified AMS:** 

**DUSTHUNTER T100 for dust** 

Manufacturer:

SICK Engineering GmbH

Bergener Ring 27

01458 Ottendorf-Okrilla

Germany

**Test Institute:** 

TÜV Rheinland Energy GmbH

This is to certify that the AMS has been tested and found to comply with the standards EN 15267-1 (2009), EN 15267-2 (2009), EN 15267-3 (2008) and EN 14181 (2015).

Certification is awarded in respect of the conditions stated in this certificate (this certificate contains 14 pages).

The present certificate replaces certificate 0000036946\_01 dated 18 July 2017.



Suitability Tested EN 15267 QAL1 Certified Regular Surveillance

www.tuv.com ID 0000036946

Publication in the German Federal Gazette (BAnz) of 20 July 2012

German Environment Agency Dessau, 20 July 2022 This certificate will expire on: 19 July 2027

TÜV Rheinland Energy GmbH Cologne, 19 July 2022

Usul y

Dr. Marcel Langner Head of Section II 4.1 D. Pat Win

ppa. Dr. Peter Wilbring

www.umwelt-tuv.eu

tre@umwelt-tuv.eu Tel. + 49 221 806-5200 TÜV Rheinland Energy GmbH Am Grauen Stein 51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

gal1.de

info@qal.de

page 1 of 14



#### Certificate:

0000036946\_02 / 20 July 2022



Test report:

936/21210076/A dated 24 October 2008

Initial certification:

20 August 2012

**Expiry date:** 

19 July 2027

Certificate:

Renewal (of previous certificate 0000036946\_01 of

18. Juli 2017 valid until 19 July 2022)

**Publication:** 

BAnz. 11 March 2009, No. 38, p. 899, chapter I no. 1.5

#### Approved application

The tested AMS is suitable for use at plants according to directive 2001/80/EC (13th BlmSchV:2009), 2000/76/EC (17th BlmSchV:2009), 30th BlmSchV:2009, Directive 2015/2193/EC (44th BlmSchV:2021), TA Luft:2002 and at plants according to the 27th BlmSchV:1997. The measured ranges have been selected so as to ensure as broad a field of application as possible.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a 6 month field test at a waste incineration plant.

The AMS is approved for an ambient temperature range of -20° to +50°C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the emission limit values relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the installation at which it will be installed.

#### Note:

The legal regulations mentioned do not correspond to the current state of legislation in every case. Each user should, if necessary, in consultation with the competent authority, ensure that this AMS meets the legal requirements for the intended use. In addition, it cannot be ruled out that legal regulations governing the use of a measuring device for emission monitoring may change during the lifetime of the certificate.

#### Basis of the certification

This certification is based on:

- Test report 936/21210076/A dated 24 October 2008 of TÜV Rheinland Immissionsschutz und Energiesysteme GmbH
- Suitability announced by the German Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process





Publication in the German Federal Gazette: BAnz. 11 March 2009, No. 38, p. 899, chapter I no. 1.5, Announcement by UBA from 19 February 2009:

#### **AMS** designation:

**DUSTHUNTER T100** 

#### Manufacturer:

SICK Engineering GmbH, Ottendorf-Okrilla

#### Field of application:

For plants requiring official approval and for plants according to the 27th BlmSchV

#### Measuring ranges during the performance test:

Dust (Transmission measurement) Certification range:

0 - 0.1 Ext.  $\triangleq 15$  mg/m³ dust at 5 m measurement path length and

0 - 0.05 Ext.

0 - 0.2 Ext.

0 - 0.5 Ext.

0 - 1.0 Ext.

#### Software versions:

MCU: 1.026, Sensor: 1.3.04,

Operating software SOPAS ET: 02.16

#### Restriction:

The measuring system shall only be employed if a fall below the dew point can be excluded

#### Notes:

- 1. A three-month period has been specified as maintenance interval.
- 2. Dust concentrations are measured in wet stack gas under operating conditions.
- 3. Complementary test to the announcement of the German Federal Environmental Agency of 12 August 2008 (BAnz. p. 3243) for the T200 measuring system.
- 4. In contrast to the T200 measuring system, T100 features no automatic deflection compensation and only a one-sided contamination measurement. With exception of these two differences, both measuring systems are identical in design.

#### **Test report:**

TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Cologne Report No.: 936/21210076/A dated 24 October 2008





Publication in the German Federal Gazette: BAnz. 26. Januar 2011, Nr. 14, S. 294, chapter IV notification 13, Announcement by UBA dated 10 January 2011:

## Notification as regards Federal Environment Agency (UBA) notice of 19 February 2009 (BAnz. p. 899, chapter I no. 1.5)

The current software versions of the dust concentration measuring system DUSTHUNTER T100 by SICK Engineering GmbH are:

MCU Firmware:

01.04.00

MCU Hardware:

1.8

Software Sensor (measuring head):

01.06.00

A notified version of the software platform SOPAS ET is necessary to ensure a full operation of the measuring system.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 5 October 2010

Publication in the German Federal Gazette: BAnz. 26. Januar 2011, Nr. 14, S. 294, chapter IV notification 30, Announcement by UBA dated 10 January 2011:

# 30 Notification to the announcement of the Federal Environmental Agency concerning suitability-tested measuring systems by SICK Engineering GmbH and SICK MAIHAK GmbH (Extract)

	Ser. no.	Measuring system/ Manufacturer	Notification	Announcement	Statement of testing body	
	5	DUSTHUNTER T100/ SICK Engineering GmbH	to announce- ment 13 of this notification	The current soft- ware version of the platform SOPAS ET for operating the measuring sys- tem is:	Energie und Umwelt GmbH of 8 November	
				SOPAS ET 2.32		





Publication in the German Federal Gazette: BAnz AT 20.07.2012 B11, chapter IV notification 17, Announcement by UBA from 06 July 2012:

17 Notification to the announcement of the Federal Environmental Agency of 19 February 2009 (BAnz. p. 899, chapter I no. 1.5) and of 10 January 2011 (BAnz. p. 294, chapter IV notification 13 and 30)

The measuring system DUSTHUNTER T100 for dust by SICK Engineering GmbH and its manufacture and quality management system fulfil the requirements of Directive EN 15267.

Concerning the application of EN 15267 to this measuring system, the following remark is added: The requirements of suitability testing according to EN 15267-3 regarding the determination coefficient R<sup>2</sup> of the calibration function were not fulfilled.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 20 March 2012

Publication in the German Federal Gazette: BAnz AT 05.03.2013 B10, chapter V notification 22, Announcement by UBA dated 12 February 2013:

Notification as regards Federal Environment Agency (UBA) notices of 19 February 2009 (BAnz. p. 899, chapter I no. 1.5) and of 6 July 2012 (BAnz AT 20.07.2012 B11, chapter IV notification 17)

The current software versions of the DUSTHUNTER T100 dust measuring system manufactured by SICK Engineering GmbH are:

MCU Firmware:

01.08.00

MCU Hardware:

1.8

Software Sensor (measuring head):

01.09.00

Statement of TÜV Rheinland Energie und Umwelt GmbH of 15 October 2012





Publication in the German Federal Gazette: BAnz AT 23.07.2013 B4, chapter V notification 13, Announcement by UBA dated 03 July 2013:

## 13 Notification as regards Federal Environmental Agency notices regarding performance tested AMS manufactured by SICK Engineering GmbH and by SICK AG (Extract)

Ser.	Measuring sys-	Notification	Announcement	Statement of
no.	tem/			testing body
	Manufacturer			
1	DUSTHUNTER	of 19 February	The current soft-	TÜV Rheinland
	T100/ SICK Engi-	2009 Federal	ware version of	Energie und
	neering GmbH	Gazette (BAnz.	the platform	Umwelt GmbH
		p. 899, chapter I	SOPAS ET for	of 25 March
		no. 1.5) and 12	operating the	2013
		February 2013	measuring sys-	
		Federal Gazette	tem is:	
		(BAnz AT	SOPAS ET 2.38	
		05.03.2013 B10,		1633
		chapter V notifi-		
		cation 22)		

Publication in the German Federal Gazette: BAnz AT 26.08.2015 B4, chapter V notification 12, Announcement by UBA dated 22 July 2015:

12 Notification as regards Federal Environment Agency (UBA) notices of 19 February 2009 (BAnz. p. 899, chapter I number 1.5) and of 3 July 2013 (BAnz AT 23.07. 2013 B4, chapter V notification 13, seq. no. 1)

The current software versions for the DUSTHUNTER T100 measuring system for dust, manufactured by SICK Engineering GmbH, are:

MCU firmware: 01.12.00 Software sensor: 1.10.02

The SOPAS ET software platform is available in a notified version for operating the AMS. The latest notified version is: SOPAS ET 2.38.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 24 March 2015





Publication in the German Federal Gazette: BAnz AT 01.08.2016 B11, chapter V notification 10, Announcement by UBA dated 14 July 2016:

10 Notification as regards Federal Environment Agency (UBA) notices of 19 February 2009 (BAnz. p. 899, chapter I number 1.5) and of 22 July 2015 (BAnz AT 26.08.2015 B4, chapter V notification 12)

The current software versions of the DUSTHUNTER T100 particle monitor manufactured by SICK Engineering GmbH are:

MCU Firmware:

01.12.02

Software Sensor:

1.10.02

For the control of the measuring system the SOPAS ET software platform is required in a notified version. The most recent notified version is: SOPAS ET 2.38

Statement issued by TÜV Rheinland Energy GmbH of 25 April 2016

Publication in the German Federal Gazette: BAnz AT 31.07.2017 B12, chapter II notification 24, Announcement by UBA dated 13 July 2017:

24 Notification as regards Federal Environment Agency (UBA) notices of 19 February 2009 (BAnz p. 899, chapter I number 1.5) and of 14 July 2016 (BAnz AT 01.08.2016 B11, chapter V, notification 10)

As an alternative light source to the XR-E LED the XM-L LED manufactured by the same manufacturer may be used for the DUSTHUNTER T100 measuring system monitoring dust manufactured by SICK Engineering GmbH. This change does not significantly affect the performance of the AMS.

Statement issued by TÜV Rheinland Energy GmbH dated 27 January 2017

Publication in the German Federal Gazette: BAnz AT 26.03.2018 B8, chapter V notification 45, Announcement by UBA dated 21 February 2018:

45 Notification as regards Federal Environment Agency (UBA) notices of 19 February 2009 (BAnz. p. 899, chapter I no. 1.5) and of 13 July 2017 (BAnz AT 31.07.2017 B12, chapter II notification 24)

The current software versions of the DUSTHUNTER T100 particle monitor for dust manufactured by SICK Engineering GmbH are as follows:

MCU:

01.12.03

Software Sensor:

1.12.00

For the control of the measuring system the SOPAS ET software platform is required in a publically notified version. The most recent publically notified version is: SOPAS ET 2.38

Statement issued by TÜV Rheinland Energy GmbH dated 28 September 2017





Publication in the German Federal Gazette: BAnz AT 17.07.2018 B9, chapter III notification 21, Announcement by UBA dated 03 July 2018:

21 Notification as regards Federal Environment Agency (UBA) notices of 19 February 2009 (BAnz. p. 899, chapter I number 1.5) and of 21 February 2018 (BAnz AT 26.03.2018 B8, chapter V notification 45)

The current software versions of the DUSTHUNTER T100 particle monitor manufactured by SICK Engineering GmbH are:

MCU:

01.12.03

Software Sensor:

1.12.02

For the control of the measuring system the SOPAS ET software platform is required in a publically notified version. The most recent publically notified version is: SOPAS ET 2.38

Statement issued by TÜV Rheinland Energy GmbH dated 2 May 2018

Publication in the German Federal Gazette: BAnz AT 22.07.2019 B8, chapter V notification 17, Announcement by UBA dated 28 June 2019:

17 Notification as regards Federal Environment Agency (UBA) notices of 19 February 2009 (BAnz. p. 899, chapter II number 1.5) and of 3 July 2018 (BAnz AT 17.07.2018 B9, chapter III notification 21)

Instead of the MCU used so far, the DUSTHUNTER T100 measuring system for dust manufactured by SICK Engineering GmbH may also be operated with the new MCU100 control unit. The latest software versions of the measuring system are:

DH T100:

1.12.02

MCU:

01.12.04

MCU100:

r2.3.6

Statement issued by TÜV Rheinland Energy GmbH dated 28 February 2019





Publication in the German Federal Gazette: BAnz AT 03.05.2021 B9, chapter III notification 49, Announcement by UBA dated 31 March 2021:

49 Notification as regards Federal Environment Agency (UBA) notices of 19 February 2009 (BAnz. p. 899, chapter I number 1.5) and of 28 June 2019 (BAnz AT 22.07.2019 B8, chapter V notification 17)

The latest software versions of the DUSTHUNTER T100 measuring system for dust manufactured by SICK Engineering GmbH are:

T100:

01.12.02

MCU:

01.12.05

MCU100:

r2.3.6

Statement issued by TÜV Rheinland Energy GmbH dated 18 September 2020

Publication in the German Federal Gazette: BAnz AT 05.08.2021 B5, chapter IV notification 41, Announcement by UBA dated 29 June 2021:

41 Notification as regards Federal Environment Agency (UBA) notices of 19 February 2009 (BAnz. p. 899, chapter I number 1.5) and of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 49)

The latest software versions of the DUSTHUNTER T100 measuring system for dust manufactured by SICK Engineering GmbH are:

T100:

01.12.03,

MCU:

01.14.00,

MCU100:

r2.3.6

Statement issued by TÜV Rheinland Energy GmbH dated 03 May 2021

#### **Certified product**

This certificate applies to automated measurement systems conforming to the following description:

The measuring system uses the measured quantities of transmission, opacity and extinction, wherein transmission is determined as the primary optical quantity. The other measured quantities are derived thereof.

The DUSTHUNTER T100 measuring system comprises the following parts in its tested device version:

- DHT-T sender/receiver unit
- Signal cable for connecting the sender/receiver unit to the control unit
- DHT-R reflector
- MCU control unit for control, evaluation and output of data
- with integrated purge air supply, for internal duct pressure of -50 ... +2 mbar
- without integrated purge air supply, in this case the following is required:
- external purge air unit, for internal duct pressure of -50 ... +30 mbar





#### Sender/Receiver unit

The sender/receiver unit contains the optical and electronic modules for sending and receiving the reflected light beam during transmission measurement. It also holds the modules for processing and evaluating signals. An integrated swivel element serves to measure contami-nation.

A purge air nozzle provides clean air for cooling the probe and avoiding contamination of the optical surfaces. The sender/receiver unit is mounted to the duct by a flange with tube.

#### Reflector

This unit contains a reflector used for redirecting the sent light beam back to the receiver in the sender/receiver unit.

#### **MCU Control unit**

The control unit has the following functions:

- Control of data traffic and processing of data from the connected unit(s)
- Signal output via analogue output (measured value) and relay outputs (device status)
- Signal input via analogue and digital inputs
- Voltage supply to the connected units
- Communication with external systems, e.g. over USB interface. In this way, the setup of plant and device parameters can be easily and comfortably carried out via laptop with the operating software. The parameters are efficiently saved in the MCU in the event of a power outage.

#### Standard interfaces

Analogue outputs:

3 outputs 0/2/4 - 22 mA (active, galvanically isolated) for output of transmission and scattered light intensity, 12 bit resolution

Relay outputs:

- 5 changeover contacts (120 V AC, 1 A, 30 V DC 2A) for output of status signals:
- Operation/Malfunction Maintenance Function check Service requirement Limit value Analogue inputs:
- 2 inputs 0 ... 20 mA (standard; without galvanic isolation) or 0 ... 5/10 V, 10 bit resolution Digital inputs:
- 4 inputs for connecting potential-free contacts (e.g. for connecting a maintenance switch or triggering control cycle)

Communication:

- USB 1.1 and RS232 (on grips) for measured value enquiry, parameterisation and soft-ware update
- RS485 for sensor connection

#### External purge air unit

In the event that the internal duct pressure exceeds +2 mbar, the integrated purge air supply can no longer be used. In this case, the "external purge air unit" shall be employed. It comprises a powerful blower and can be used at overpressures in the duct of up to 30 mbar. The MCU with integrated purge air supply (MCU-P) was used during the field test.





#### Auxiliary equipment for device testing

Testing materials for linearity test

The linearity test serves for checking the correct functioning of transmission measurement (see service manual). In order to do this, filter glasses with predetermined transmission values are inserted in the beam path. The values of the filter glasses are subsequently compared with those measured by DUSTHUNTER T100. If they are within the permissible tolerance range, it means that the measuring system works correctly.

#### Adjustment mounts for normalisation

In order to test measurements on smoke-free measurement sections, adjustment mounts with built-on sender/receiver and reflector units are used. These are placed at a defined distance from one another and directed in a way that the optical axes coincide. The thereby de-termined transmission value is set to 100 %, providing the standard for measurement in the section loaded with dust.

#### Waste gas free section for normalisation

Instead of adjustment mounts, the measuring system can be normalised inserting a tube with the exact length of the duct between sender/receiver and reflector units, thus creating a waste gas free measurement section. The assembling and alignment of the sender/receiver and reflector units on a smoke-free section becomes thereby easy and precise. The tube can be sealed off with end caps when not at use, in order to avoid infiltration of dust. The tube is especially recommended if a dust-free environment cannot be provided for normalisation.

#### **General notes**

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document and the certification mark remains property of TÜV Rheinland Energy GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: **qal1.de**.





#### **History of documents**

Certification of DUSTHUNTER T100 is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

#### **Basic test**

Test report 936/21207351/C dated 10 March 2008
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH
Publication BAnz. 03 September 2008, No. 133, p. 3243, chapter I number 1.2
UBA announcement dated 12 August 2008

#### Supplementary testing

Test report 936/21210076/A dated 24 October 2008
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH
Publication BAnz. 11 March 2009, No. 38, p. 899, chapter I number 1.5
UBA announcement dated 19 February 2009

#### **Notifications**

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 5 October 2010 Publication BAnz. 26 January 2011, No. 14, p. 294, chapter IV notification 13 UBA announcement dated 10 January 2011 (Software changes)

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 8 November 2010 Publication BAnz. 26 January 2011, No. 14, p. 294, chapter IV notification 30 UBA announcement dated 10 January 2011 (Software changes)

#### Initial certification according to EN 15267

Certificate No. 0000036946\_00: 20 August 2012
Expiry date of the certificate: 19 July 2017
Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 20 March 2012
Test report 936/21210076/A dated 24 October 2008
Publication BAnz AT 20.07.2012 B11, chapter IV number 17
UBA announcement dated 6 July 2012

#### **Notifications**

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 15 October 2012 Publication BAnz AT 05.03.2013 B10, chapter V notification 22 UBA announcement dated 12 February 2013 (Software changes)

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 25 March 2013 Publication BAnz AT 23.07.2013 B4, chapter V notification 13 UBA announcement dated 3 July 2013 (Software changes)

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 24 March 2015 Publication BAnz AT 26.08.2015 B4, chapter V notification 12 UBA announcement dated 22 July 2015 (Software changes)



#### Certificate:

0000036946\_02 / 20 July 2022



Statement issued by TÜV Rheinland Energy GmbH dated 25 April 2016 Publication BAnz AT 01.08.2016 B11, chapter V notification 10 UBA announcement dated 14 July 2016 (Software changes)

#### Renewal of certificate

Certificate No. 0000036946\_01: 18 July 2017 Expiry date of the certificate: 19 July 2022

#### **Notifications**

Statement issued by TÜV Rheinland Energy GmbH dated 27 January 2017 Publication BAnz AT 31.07.2017 B12, chapter II notification 24 UBA announcement dated 13 July 2017 (Hardware changes)

Statement issued by TÜV Rheinland Energy GmbH dated 28 September 2017 Publication BAnz AT 26.03.2018 B8, chapter V notification 45 UBA announcement dated 21 February 2018 (Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 2 May 2018 Publication BAnz AT 17.07.2018 B9, chapter III notification 21 UBA announcement dated 3 July 2018 (Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 28 February 2019 Publication BAnz AT 22.07.2019 B8, chapter V notification 17 UBA announcement dated 28 June 2019 (Soft- and hardware changes)

Statement issued by TÜV Rheinland Energy GmbH dated 18 September 2020 Publication BAnz AT 03.05.2021 B9, chapter III notification 49 UBA announcement dated 31 March 2021 (Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 3 May 2021 Publication BAnz AT 05.08.2021 B5, chapter IV notification 41 UBA announcement dated 29 June 2021 (Software changeSoftwareänderung)

#### Renewal of certificate

Certificate No. 0000036946\_02: 20 July 2022 Expiry date of the certificate: 19 July 2027





#### EN ISO 14956 and EN 15267-3 calculation for QAL1 in EN 14181

Ma	nuta	icture	er data	

Manufacturer Name of measuring system Serial Number Measuring Principle SICK Enginnering GmbH DUSTHUNTER T100 Geräte 03 /04/ 05 /06 Transmissionsmessung

#### **TÜV Data**

Approval Report Date Editor 936/21210076/A 24.10.2008 Röllig

#### **Measurement Component**

certificated range

Staub

15 mg/m<sup>3</sup>

#### Calculation of the combined standard uncertainty

Test Value		$\Delta X_{max, j}$	U <sup>2</sup>
Standard deviation from paired measurements under field cond	0.21 mg/m <sup>3</sup>	0.045	
Lack of fit	$u_{d.z}$	0.15 mg/m <sup>3</sup>	0.008
Zero drift from field test	$u_{d.s}$	0.06 mg/m <sup>3</sup>	0.001
Span drift from field test	Ut	-0.39 mg/m <sup>3</sup>	0.051
Influence of ambient temperature at span	u <sub>n</sub>	0.02 mg/m <sup>3</sup>	0.000
Influence of supply voltage	Uf	-0.12 mg/m <sup>3</sup>	0.005
Influence of sample pressure	Ui	0.00 mg/m <sup>3</sup>	0.000
Uncertainty of reference material	U <sub>rm</sub>	0.30 mg/m <sup>3</sup>	0.030
Excursion of measurement beam	$u_{mb}$	0.30 mg/m <sup>3</sup>	0.030

<sup>\*</sup> The greater value of: "Repeatability standard deviation at span" or "Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u <sub>C</sub> )	$u_c = \sqrt{\sum (u_{max, j})^2}$	0.411
Total expanded uncertainty	$U = u_c * k = u_c * 1,96$	0.805
Relative total expanded uncertainty	U in % of the ELV 10 mg/m <sup>3</sup>	8.1
Requirement	U in % of the ELV 10 mg/m <sup>3</sup>	22.5