

# CONFIRMATION

Notification: 0000001012\_02\_01  
on changes according to EN 15267 regarding certificate 0000001012\_02 dated 22 July 2016

---

**Measuring system:** GMS810-FIDORi

**Manufacturer:** SICK AG  
Nimburger Straße 11  
79276 Reute  
Germany

---

**German Federal Environmental Agency (UBA)**

**Announcement about the uniform practice in  
monitoring emissions and ambient air  
dated 14 July 2016  
Federal Gazette: BAnz AT 01.08.2016 B11**

**V. Notifications to the uniform practice for the continuous monitoring  
of emission and ambient air:**

**18 Notification as regards Federal Environmental Agency (UBA) notices  
of 23 February 2012 (BAnz. p. 920, chapter I number 2.1) and  
of 18 February 2016 (BAnz AT 14.03.2016 B7, chapter V notification 41)**


The maintenance interval for the GMS810-FIDORi measuring system for total organic carbon  
manufactured by SICK AG is three months.

Test results are detailed in test report number 936/21229847/C of 19 February 2016 issued  
by TÜV Rheinland Energie und Umwelt GmbH.


Statement issued by TÜV Rheinland Energy GmbH dated 9 May 2016.

---

TÜV Rheinland Energy GmbH  
Cologne, 18 August 2016



i. V. Dipl.-Ing. Guido Baum



i. A. Dipl. Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.