

CONFIRMATION

Notification: 0000001012_02_02
on changes according to EN 15267 regarding certificate 0000001012_02 dated 22 July 2016

Measuring system: GMS810-FIDOR for TOC

Manufacturer: SICK AG
Poppenbütteler Bogen 9b
22399 Hamburg
Germany

German Federal Environmental Agency (UBA)

**Announcement about the uniform practice in
monitoring emissions and ambient air
dated 22 February 2017
Federal Gazette: BAnz AT 15.03.2017 B6**

V Notifications to the uniform practice for the continuous monitoring of emission and ambient air:

- 26 Notification as regards Federal Environment Agency (UBA) notices
of 23 February 2012 (BAnz. p. 920, chapter I number 2.1) and
of 14 July 2016 (BAnz AT 01.08.2016 B11, chapter V, 18th notification)


The BCU of the modular GMS810-FIDOR/GMS 810 FIDORi measuring system for total-C manufactured by SICK AG now provides a digital Modbus interface (RTU and TCP/IP) in accordance with VDI guideline 4201, parts 1 and 3. Test results are detailed in test report number 936/21236082/B of 10 October 2016 issued by TÜV Rheinland Energy GmbH.


The current BCU software version is 9150883_4.003 Aug 22 2016 1449.

The GMS 810 FIDOR/GMS 810 FIDORi measuring system may also be operated with the system housing for wall mounting. When used with the housing for wall mounting, the measuring system is called GMS 840-FIDOR/GMS-840 FIDORi.

Statement issued by TÜV Rheinland Energy GmbH dated 10 October 2016

TÜV Rheinland Energy GmbH
Cologne, 28 March 2017


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl. Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.