

# CONFIRMATION

Notification: 0000025930\_03\_02  
on changes according to EN 15267 regarding certificate 0000025930\_03 dated 02 February 2015

---

**Measuring system:** PowerCems100 for CO, NO, NO<sub>2</sub>, SO<sub>2</sub>, CH<sub>4</sub>, N<sub>2</sub>O, CO<sub>2</sub> and O<sub>2</sub>

**Manufacturer:** SICK AG  
Nimburger Str. 11  
79276 Reute  
Germany

---

## German Federal Environmental Agency (UBA)

**Announcement about the uniform practice in  
monitoring emissions and ambient air  
dated 22 February 2017  
Federal Gazette: BAnz AT 15.03.2017 B6**

**V Notifications to the uniform practice for the continuous monitoring  
of emission and ambient air:**

- 25 Notification as regards Federal Environment Agency (UBA) notices  
of 23 February 2012 (BAnz. p. 920, chapter I number 5.1) and  
of 18 February 2016 (BAnz AT 14.03.2016 B7, chapter V, 34<sup>th</sup> notification)


The BCU of the modular PowerCEMS100 measuring system for CO, NO, NO<sub>2</sub>, SO<sub>2</sub>,  
CH<sub>4</sub>, N<sub>2</sub>O, CO<sub>2</sub> and O<sub>2</sub> manufactured by SICK AG now provides a digital Modbus  
interface (RTU and TCP/IP) in accordance with VDI guideline 4201, parts 1 and 3.  
Test results are detailed in test report number 936/21236082/A of 10 October 2016  
issued by TÜV Rheinland Energy GmbH.


The current BCU software version is 9150883\_4.003 Aug 22 2016 1449.

Statement issued by TÜV Rheinland Energy GmbH dated 10 October 2016

---

TÜV Rheinland Energy GmbH  
Cologne, 28 March 2017

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl. Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.