

CONFIRMATION

Notification: 0000032301_01_02
on changes according to EN 15267 regarding certificate 0000032301_01 dated 05 March 2018

AMS designation: PG 350 E for CO, NOx, SO₂, O₂ und CO₂

Manufacturer: HORIBA Europe GmbH
Julius-Kronenberg-Str. 9
42799 Leichlingen
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 27 May 2020**

Publication: Federal Gazette: BAnz AT 31.07.2020 B10


**II. Notifications as regards the uniform practice in monitoring
emissions and ambient air**

**12 Notification as regards Federal Environment Agency (UBA) notices
of 12 February 2013 (BAnz AT 05.03.2013 B10, chapter I number 5.2) and
of 21 February 2018 (BAnz AT 26.03.2018 B8, chapter V, notification 28)**

FINEFLEX BIO™ Board TOMBO No. 5625 may now be used as a material for thermal insulation of the NOx converter used for the PG-350E measuring system for NOx, SO₂, CO, CO₂ and O₂ manufactured by HORIBA Europe GmbH. The insulation material may also be used for instrument version PG-350EDR.

Statement issued by TÜV Rheinland Energy GmbH dated 10 March 2020

TÜV Rheinland Energy GmbH
Cologne, 07 August 2020


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.