

CONFIRMATION

Notification: 0000035015_03_03
on changes according to EN 15267 regarding Certificate no. 0000035015_03 dated 28 February 2017

AMS designation: **MERC300Z and MERC300Z Indoor for Hg**

Manufacturer: **SICK AG**
 Nimburger Straße 11
 79276 Reute
 Germany

German Federal Environment Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 3 July 2018
Publication: BAnz AT 17.07.2018 B9**

III. Notifications as regards the uniform practice for the continuous monitoring of emissions and ambient air:

24 Notification as regards Federal Environment Agency (UBA) notices dated 12 February 2013 (BAnz AT 05.03.2013 B10, chapter I number 2.3) and dated 13 July 2017 (BAnz AT 31.07.2017 B12, chapter II 23rd notification)

As of software version 9191789_ZS08, the MERC300Z and MERC300Z Indoor measuring systems manufactured by SICK AG provide a digital Modbus interface (TCP/IP) in accordance with VDI guideline 4201, parts 1 and 3.

Report no. 936/21242227/B dated 2 May 2018 issued by TÜV Rheinland Energy GmbH presents the corresponding test results.


The current software version is 9191789_ZS08.


The following software versions may also be used:

9191789_YXF0 and 9159349_YZE7.

Statement issued by TÜV Rheinland Energy GmbH dated 2 May 2018

TÜV Rheinland Energy GmbH
Cologne, 1 August 2018


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Phone: +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.