

CONFIRMATION

Notification: 0000036943_02_03
on changes according to EN 15267 regarding certificate 0000036943_02 of 18 July 2017

AMS designation: DUSTHUNTER SB100 for dust

Manufacturer: SICK Engineering GmbH
Bergener Ring 27
01458 Ottendorf-Okrilla
Germany

German Federal Environment Agency (UBA)

**Announcement about the uniform practice in
monitoring emissions and ambient air
dated 28 June 2019
Publication: BAnz AT 22.07.2019 B8**

V. Notification as regards the Uniform Practice in monitoring emissions:


- 15 Notification as regards Federal Environment Agency notices
of 12 February 2013 (BAnz AT 05.03.2013 B10, chapter I number 1.6) and
of 3 July 2018 (BAnz AT 17.07.2018 B9, chapter III 26th notification)**


Instead of the MCU used so far, the DUSTHUNTER SB100 measuring system for dust
manufactured by SICK Engineering GmbH may also be operated
with the new MCU100 control unit. The latest software versions of the measuring system are:

SB 100: 1.06.03
MCU: 01.12.04
MCU100: r2.3.6

Statement issued by TÜV Rheinland Energy GmbH dated 28 February 2019

TÜV Rheinland Energy GmbH
Cologne, 15 September 2019


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Phone. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.