

# CONFIRMATION

Notification: 0000036946\_01\_04  
on changes according to EN 15267 regarding certificate 0000036946\_01 of 18 July 2017

---

**AMS designation:** DUSTHUNTER T100 for dust

**Manufacturer:** SICK Engineering GmbH  
Bergener Ring 27  
01458 Ottendorf-Okrilla  
Germany

---

**German Federal Environment Agency (UBA)**

**Announcement about the uniform practice in  
monitoring emissions and ambient air  
dated 28 June 2019  
Publication: BAnz AT 22.07.2019 B8**

**V. Notification as regards the Uniform Practice in monitoring emissions:**

- 17 Notification as regards Federal Environment Agency (UBA) notices  
of 19 February 2009 (BAnz. p. 899, chapter II number 1.5) and  
of 3 July 2018 (BAnz AT 17.07.2018 B9, chapter III 21<sup>st</sup> notification)**


Instead of the MCU used so far, the DUSTHUNTER T100 measuring system for dust  
manufactured by SICK Engineering GmbH may also be operated with the new MCU100  
control unit. The latest software versions of the measuring system are:


DH T100: 1.12.02  
MCU: 01.12.04  
MCU100: r2.3.6

Statement issued by TÜV Rheinland Energy GmbH dated 28 February 2019

---

TÜV Rheinland Energy GmbH  
Cologne, 15 September 2019

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Phone. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.