

CONFIRMATION

Notification: 0000037052_01_01
on changes according to EN 15267 regarding certificate 0000037052_01 dated 18 July 2017

AMS designation: GM32 in-situ gas analyser, GMP probe version for NO and SO₂

Manufacturer: SICK AG
Nimburger Str. 11
76276 Reute
Germany

German Federal Environment Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 13 July 2017
Publication: BAnz AT 31.07.2017 B12**


**II. Notifications as regards the uniform practice for the continuous monitoring
of emission and ambient air:**


**18 Notification as regards Federal Environment Agency (UBA) notices
of 3 August 2009 (BAnz p. 2929, chapter I number 3.3) and
of 22 February 2017 (BAnz AT 15.03.2017 B6, chapter V 28th notification)**

Deuterium lamp D2Plus-DS Type XD6201-01 (transmittance version) may now also be used
for the GM 32 in-situ gas analyser in its GMP probe version measuring NO and SO₂
manufactured by SICK AG.

Statement issued by TÜV Rheinland Energy GmbH dated 8 March 2017

TÜV Rheinland Energy GmbH
Cologne, 08 September 2017


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.