

# CONFIRMATION

Notification: 0000039317\_00\_01\_rev1  
on changes according to EN 15267 regarding certificate 0000039317 dated 20 August 2013

---

**Measuring system:** APDA-371 with PM<sub>10</sub>-pre-separator for particulate matter PM<sub>10</sub>

**Manufacturer:** HORIBA Europe GmbH  
Hans-Mess-Str. 6  
61440 Oberursel/Ts.  
Germany

---

## German Federal Environmental Agency (UBA)

Announcement about the uniform practice in  
monitoring emissions and ambient air.  
25 February 2015  
Federal Gazette BAnz AT 02 April 2015 B5

**IV. Notifications to the uniform practice for the continuous monitoring of  
emission and ambient air:**


**9 Notification as regards Federal Environment Agency (UBA) notice of  
03 July 2013 (Federal Gazette (BAnz) AT 23 July 2013 B4, chapter III number 3.1)**


The 970603 pressure sensor (MICROSWITCH #185PC15AT) of the APDA-371 measuring system with PM<sub>10</sub> pre-separator, manufactured by HORIBA Europe GmbH, has been discontinued and was replaced by the 970595 pressure sensor (HONEYWELL SSCDANN015PAAA5).

Statement of TÜV Rheinland Energie und Umwelt GmbH of 20 September 2014

---

TÜV Rheinland Energie und Umwelt GmbH  
Cologne, 30. April 2015

  
i. A. Dipl.-Ing. Ruth Steinhagen-Pinnow

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.de](http://www.umwelt-tuv.de)  
teu@umwelt-tuv.de  
Tel. +49 221 806-5200

TÜV Rheinland Energie und Umwelt GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body. This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.