

CONFIRMATION

Notification: 0000039320_00_03_rev1
on changes according to EN 15267 regarding certificate 0000039320 dated 20 August 2013

Measuring system: Modell 49i for O₃

Manufacturer: Thermo Fisher Scientific
27 Forge Parkway
Franklin, MA 02038
USA

German Federal Environmental Agency (UBA)

Announcement about the uniform practice in
monitoring emissions and ambient air.
25 February 2015
Federal Gazette BAnz AT 02 April 2015 B5

**IV. Notifications to the uniform practice for the continuous monitoring of
emission and ambient air:**

**19 Notification as regards Federal Environment Agency (UBA) notices of
21 February 2006 (Federal Gazette (BAnz.) p. 2653, chapter IV number 3.2) and of
17 July 2014 (Federal Gazette (BAnz) AT 05 August 2014 B11, chapter V 23th notification)**


The previous magnetic valves (SV-111747-501) of the Model 49i measuring system for O₃,
manufactured by Thermo Fisher Scientific, are replaced by new magnetic valves (KL344-T-1S1-
C204).

The current software version is:

V 02.02.01 (114620-00)

Statement of TÜV Rheinland Energie und Umwelt GmbH of 22 September 2014

TÜV Rheinland Energie und Umwelt GmbH
Cologne, 30. April 2015


i. A. Dipl.-Ing. Ruth Steinhagen-Pinnow


i. A. Dipl. Ing. Carsten Röllig

www.umwelt-tuv.de
teu@umwelt-tuv.de
Tel. +49 221 806-5200

TÜV Rheinland Energie und Umwelt GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body. This accreditation is
limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.