

CONFIRMATION

Notification: 0000039321_00_02
on changes according to EN 15267 regarding certificate 0000039321_00 dated 29 April 2014

Measuring system: MGA12 measuring system monitoring CO, NO, SO₂, and O₂

Manufacturer: Dr. Födisch Umweltmesstechnik AG
Zwemkauer Straße 159
04420 Markranstädt
Germany

German Federal Environmental Agency (UBA)

**Announcement about the uniform practice in
monitoring emissions and ambient air
dated 14 July 2016
Federal Gazette: BAnz AT 01.08.2016 B11**


**V. Notifications to the uniform practice for the continuous monitoring
of emission and ambient air:**


**5 Notification as regards Federal Environmental Agency (UBA) notices
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter I number 3.4) and
of 22 July 2015 (BAnz AT 26.08.2015 B4 chapter V notification 31)**

The GCU12 test gas cooler of the MGA12 measuring system monitoring CO, NO, SO₂, and O₂ manufactured by Dr. Födisch Umweltmesstechnik AG was equipped with new electronics and a new display unit. The name of the new test gas cooler is GCU16 (as of serial number 17xxx). This version serves as an alternative to its predecessor.

Statement issued by TÜV Rheinland Energy GmbH dated 27 April 2016.

TÜV Rheinland Energy GmbH
Cologne, 18 August 2016


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl. Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.