

# CONFIRMATION

Notification: 0000040201\_02\_01  
on changes according to EN 15267 regarding certificate 0000040201\_02 dated 30 September 2015

---

**AMS designation:** D-R 320 for dust

**Manufacturer:** DURAG GmbH  
Kollaustraße 105  
22453 Hamburg  
Germany

---

**German Federal Environment Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 21 February 2018  
Publication: BAnz AT 26.03.2018 B8**

**V. Notifications as regards the uniform practice for the continuous monitoring  
of emission and ambient air:**

**24 Notification as regards Federal Environment Agency notice  
of 22 July 2015 (BAnz AT 26.08.2015 B4, chapter I number 1.1)**

The current software versions of the D-R 320 measuring system for dust  
manufactured by DURAG GmbH are as follows:

D-R 320: 01.10.R0001  
D-ISC 100: 01.04R0007  
D-ESI 100: 01.10R0007

Versions of approved intermediate versions are as follows:


D-ISC 100: 01.04R0001; 01.04R0004; 01.04R0006  
D-ESI 100: 1.1.016; 1.1.017; 1.2.003


In addition to the existing XPPower DNR240PS24-I power supply, the Phoenix  
Contact QUINT4-PS/1AC/24DC/10 power supply may be used for the D-ISC 100.

Statement issued by TÜV Rheinland Energy GmbH dated 8 December 2017

---

TÜV Rheinland Energy GmbH  
Cologne, 4 June 2018

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Phone. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.