

# CONFIRMATION

Notification: 0000040213\_00\_03  
on changes according to EN 15267 regarding certificate 0000040213\_00 dated 29 April 2014

---

**Measuring system:** 5014i Beta with PM<sub>2.5</sub>- pre-separator for floating-dust PM<sub>2.5</sub>

**Manufacturer:** Thermo Fisher Scientific  
27, Forge Parkway  
Franklin, MA 02038  
USA

---

**German Federal Environmental Agency (UBA)**

**Announcement about the uniform practice in  
monitoring emissions and ambient air  
dated 14 March 2016  
Federal Gazette: BAnz AT 14.03.2016 B7**


- V. **Notifications to the uniform practice for the continuous monitoring of emission and ambient air:**
- 16 **Notification as regards Federal Environment Agency (UBA) notices of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 6.2) and of 25 February 2015 (BAnz AT 02.04.2015 B5, chapter IV notification 23)**

The 5014i Beta measuring system with PM<sub>2.5</sub> pre-separator for suspended particulate matter PM<sub>2.5</sub> manufactured by Thermo Fisher Scientific can be operated with the vacuum pump type GAST 87R647-PDS-HV-913 also.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 22 October 2015

---

TÜV Rheinland Energy GmbH  
Cologne, 26 April 2016

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl. Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.