

CONFIRMATION

Notification: 0000040215_00_02_rev1
on changes according to EN 15267 regarding certificate 0000040215 dated 29 April 2014

Measuring system: Modell 5030i SHARP with PM_{2.5}-pre-separator
for particulate matter PM_{2.5}

Manufacturer: Thermo Fisher Scientific
27 Forge Parkway
Franklin, MA 02038
USA

German Federal Environmental Agency (UBA)

Announcement about the uniform practice in
monitoring emissions and ambient air.
25 February 2015
Federal Gazette BAnz AT 02 April 2015 B5

**IV. Notifications to the uniform practice for the continuous monitoring of
emission and ambient air:**

**25 Notification as regards Federal Environment Agency (UBA) notices of 27 February 2014
(Federal Gazette (BAnz) AT 01 April 2014 B12, chapter IV number 6.3) and of 17 July 2014
(Federal Gazette (BAnz) AT 05 August 2014 B11, chapter V 26th notification)**


The current software version for the Model 5030i SHARP measuring system with PM_{2.5} pre-separator for the suspended particulate matter PM_{2.5}, manufactured by Thermo Fisher Scientific, is:

V 02.02.05 (111578-00).

In the future the valve for the automatic zero point equalization will have a nickel-plated housing and will be equipped with a Viton Elastomer seal.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 22 September 2014

TÜV Rheinland Energie und Umwelt GmbH
Cologne, 30. April 2015


i. A. Dipl.-Ing. Ruth Steinhausen-Pinnow


i. A. Dipl. Ing. Carsten Röllig

www.umwelt-tuv.de
teu@umwelt-tuv.de
Tel. +49 221 806-5200

TÜV Rheinland Energie und Umwelt GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body. This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.