

CONFIRMATION

Notification: 0000040215_00_03
on changes according to EN 15267 regarding certificate 0000040215_00 dated 29 April 2014

Measuring system: 5030i SHARP with PM_{2.5}- pre-separator for floating-dust PM_{2.5}

Manufacturer: Thermo Fisher Scientific
27, Forge Parkway
Franklin, MA 02038
USA

German Federal Environmental Agency (UBA)


**Announcement about the uniform practice in
monitoring emissions and ambient air
dated 14 March 2016
Federal Gazette: BAnz AT 14.03.2016 B7**

- V. Notifications to the uniform practice for the continuous monitoring of
emission and ambient air:**
- 17 Notification as regards Federal Environment Agency (UBA) notices of
27 February 2014 (BAnz AT 01.04.2014 B12, chapter IV number 6.3) and of
25 February 2015 (BAnz AT 02.04.2015 B5, chapter IV notification 25)**

The 5030i SHARP measuring system with PM_{2.5}- pre-separator for suspended particulate matter PM_{2.5} manufactured by Thermo Fisher Scientific can also be operated with the vacuum pump type GAST 87R647-PDS-HV-913.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 22 October 2015

TÜV Rheinland Energy GmbH
Cologne, 26 April 2016


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl. Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.