

CONFIRMATION

Notification: 0000040334_00_02
on changes according to EN 15267 regarding certificate 0000040334_00 dated 9 September 2014

Measuring system: MEAC2012

Manufacturer: SICK AG
Nimburger Str. 11
79276 Reute
Germany

German Federal Environmental Agency (UBA)

**Announcement about the uniform practice in
monitoring emissions and ambient air
dated 14 March 2016
Federal Gazette: BAnz AT 14.03.2016 B7**


- V. **Notifications to the uniform practice for the continuous monitoring of
emission and ambient air:**
- 19 **Notification as regards Federal Environment Agency (UBA) notices of
12 February 2013 (BAnz AT 05.03.2013 B10, chapter III number 1.1) and of
22 July 2015 (BAnz AT 26.08.2015 B11, chapter V notification 1)**


The current software version of the data evaluating system MEAC2012
from SICK AG, Hamburg is:
Version 3.6

In plants with tele data transmission a software update is necessary. The transfer of
messages and identifier of measured values were corrected.
In plants according to the 17th and/or 30th BImSchV, a software update to the latest
version is recommended. The message „ARE-Ausfall > 60 h” were corrected here.

Statement of TÜV Rheinland Energie und Umwelt GmbH of 14 October 2015

TÜV Rheinland Energy GmbH
Cologne, 16 April 2016


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl. Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.