

# CONFIRMATION

Notification: 0000040337\_00\_02  
on changes according to EN 15267 regarding certificate 0000040337\_00 dated 9 September 2014

---

**AMS designation:** F-701-20 with PM<sub>2.5</sub> pre-separator for suspended particles PM<sub>2.5</sub>  
**Manufacturer:** DURAG GmbH  
Kollastraße 105  
22453 Hamburg  
Germany

---

## German Federal Environment Agency (UBA)

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 21 February 2018  
Publication: BAnz AT 26.03.2018 B8**

**V. Notifications as regards the uniform practice for the continuous monitoring  
of emission and ambient air:**

**4 Notification as regards Federal Environment Agency notices  
of 17 July 2014 (BAnz AT 05.08.2014 B11, chapter III number 3.1) and  
of 22 February 2017 (BAnz AT 15.03.2017 B6, chapter V 3<sup>rd</sup> notification)**


The current software version of the F-701-20 with PM<sub>2.5</sub> pre-separator for PM<sub>2.5</sub>  
suspended particulate matter manufactured by DURAG GmbH is: 03.11R0008.

Instead of the Buschjost 8288200.9624.02400 control valve, the measuring system  
may also be equipped with the Buschjost 8288200.9638.02400 valve.

Statement issued by TÜV Rheinland Energy GmbH dated 29 September 2017

---

TÜV Rheinland Energy GmbH  
Cologne, 4 June 2018

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Phone. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.