

CONFIRMATION

Notification: 0000040337_00_03
on changes according to EN 15267 regarding Certificate no. 0000040337_00 of 9 September 2014

AMS designation: F-701-20 with PM_{2.5} pre-separator for suspended particles PM_{2.5}

Manufacturer: DURAG GmbH
Kollastr. 105
22453 Hamburg
Germany

German Federal Environment Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 3 July 2018
Publication: BAnz AT 17.07.2018 B9**

III. Notifications as regards the uniform practice for the continuous monitoring of emissions and ambient air:

- 27 Notification as regards Federal Environment Agency (UBA) notices dated 17 July 2014 (BAnz AT 05.08.2014 B11, chapter III number 3.1) and dated 21 February 2018 (BAnz AT 26.03.2018 B8, chapter V 4th notification)**

The current software version of the F-701-20 measuring system with PM_{2.5} pre-separator for suspended particulate matter PM_{2.5} manufactured by DURAG GmbH is:


04.11R0009


The housing of the measuring system was adapted to be mounted in a 19" measuring rack.

The instrument is also available as a version with external pump. The letter "F" in the model code pattern F-701-20 PM xx2-xxxxxF clearly marks and identifies this instrument version.

Statement issued by TÜV Rheinland Energy GmbH dated 2 May 2018

TÜV Rheinland Energy GmbH
Cologne, 1 August 2018


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Phone: +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.