

# CONFIRMATION

Notification: 0000043107\_01\_01  
on changes according to EN 15267 regarding certificate 0000043107\_01 dated 25 April 2016

---

**Measuring system:** APDA 372 for PM<sub>10</sub> and PM<sub>2,5</sub>

**Manufacturer:** HORIBA Europe GmbH  
Hans-Mess-Str. 6  
61440 Oberursel / Ts.  
Germany

---

**German Federal Environmental Agency (UBA)**

**Announcement about the uniform practice in  
monitoring emissions and ambient air  
dated 14 July 2016  
Federal Gazette: BAnz AT 01.02.2016 B11**

**V. Notifications to the uniform practice for the continuous monitoring  
of emission and ambient air:**

**34 Notification as regards Federal Environmental Agency (UBA) notices  
of 25 February 2015 (BAnz AT 02.04.2015 B5, chapter III number 3.1) and  
of 18 February 2016 (BAnz AT 14.03.2016 B7 chapter V notification 4)**

The sensitivity test of the particle sensor for the APDA 372 PM<sub>10</sub> and PM<sub>2,5</sub> particle monitor with MonoDust 1500 manufactured by Horiba Europe GmbH can be performed at an IADS temperature between 35 °C and 50 °C.


The measuring system may provide two additional contacts for the control of an external pump/flow regulator (not relevant for the performance-tested instrument version).

The current software version of the measuring system is:  
100408.0014.0001.0001.0011

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 24 February 2016.

---

TÜV Rheinland Energy GmbH  
Cologne, 18 August 2016

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl. Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.