

CONFIRMATION

Notification: 0000043107_02_02
on changes according to EN 15267 regarding certificate 0000043107_02 of 12 June 2019

AMS designation: APDA-372 and APDA-372E for PM₁₀ and PM_{2.5}

Manufacturer: HORIBA Europe GmbH
Hans-Mess-Str. 6
61440 Oberursel/Ts.
Germany

German Federal Environment Agency (UBA)

**Announcement about the uniform practice in
monitoring emissions and ambient air
dated 24 February 2020
Publication: BAnz AT 24.03.2020 B7**

**IV. Notification as regards the Uniform Practice
in monitoring emissions and ambient air:**

**52 Notification as regards Federal Environment Agency (UBA) notices
of 25 February 2015 (BAnz AT 02.04.2015 B5, chapter III number 3.1) and
of 28 June 2019 (BAnz AT 22.07.2019 B8, chapter V 13th notification)**

The software of the APDA-372 and APDA-372E for PM₁₀ and PM_{2.5} manufactured
by HORIBA Europe GmbH has been updated.


The current software version is:
100454.0014.0001.0001.0011.


The sensitivity of the particle sensor has to be checked once every three months
using CalDust 1100 or MonoDust 1500.

The new sensor housing may be used for the APDA-372E in the future.

Statement issued by TÜV Rheinland Energy GmbH dated 19 September 2019

TÜV Rheinland Energy GmbH
Cologne, 23 April 2020


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Phone. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to certificate D-PL-11120-02-00.