

CONFIRMATION

Notification: 0000043528_00_02
on changes according to EN 15267 regarding certificate 0000043528_00 dated 30 September 2015

AMS designation: F-701-20 for suspended particles PM₁₀

Manufacturer: DURAG GmbH
Kollaustraße 105
22453 Hamburg
Germany

German Federal Environment Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 21 February 2018
Publication: BAnz AT 26.03.2018 B8**

**V. Notifications as regards the uniform practice for the continuous monitoring
of emission and ambient air:**


**5 Notification as regards Federal Environment Agency (UBA) notice
of 12 September 2006 (BAnz p. 6715, chapter IV number 1.1) and
of 22 February 2017 (BAnz AT 15.03.2017 B6, chapter V 4th notification)**


The current software version of the F-701-20 measuring system for PM₁₀ suspended
particulate matter manufactured by DURAG GmbH is: 03.11R0008.

Instead of the Buschjost 8288200.9624.02400 control valve, the measuring system
may also be equipped with the Buschjost 8288200.9638.02400 valve.

Statement issued by TÜV Rheinland Energy GmbH dated 29 September 2017

TÜV Rheinland Energy GmbH
Cologne, 4 June 2018


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Phone. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.