

# CONFIRMATION

Notification: 0000051687\_00\_01  
on changes according to EN 15267 regarding certificate 0000051687\_00 dated 19 August 2016

---

**Measuring system:** CEM-DAS

**Manufacturer:** ABB Automation GmbH  
Stierstädter Str. 5  
60488 Frankfurt/Main  
Germany

---

**German Federal Environmental Agency (UBA)**

**Announcement about the uniform practice in  
monitoring emissions and ambient air  
dated 22 February 2017  
Federal Gazette: BAnz AT 15.03.2017 B6**

**IV Corrections to the uniform practice for the continuous monitoring  
of emission and ambient air:**

- 2 Correction for the announcement of the Federal Environment Agency (UBA)  
dated 14 July 2016 (BAnz AT 01.08.2016 B11, chapter II number 1.1)

In the above-mentioned announcement of emission calculator CEM-DAS  
manufactured by ABB Automation GmbH is the company name of the manufacturer  
not correct and must say as follows:

ABB Automation GmbH (instead of ABB Automation GmbH Process Automation).

Statement issued by TÜV Rheinland Energy GmbH dated 10 October 2016

---

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

**V Notifications to the uniform practice for the continuous monitoring of emission and ambient air:**

- 12 Notification as regards Federal Environment Agency (UBA) notice of 14 July 2016 (BAnz AT 01.08.2016 B11, chapter II number 1.1)

Software implemented in the CEM-DAS emission calculator manufactured by ABB Automation GmbH has been re-engineered. In this context, the data simulation feature has also been debugged. This is why CEM-DAS version 1.2.0 may no longer be used.


The formula evaluation MAX() and the logical operators (AND) and (OR) had been performed incorrectly by the DAA software module. This correction made it necessary to install the new DAA 1.2 (003) version for parameterisations which use formulas with the relevant operators.


The current software versions of the CEM-DAS emission calculator are:

Data evaluation:	CEM-DAS:	1.2.3
Oracle data base:		11.2
Data collection:	DAA	1.2 (003)

Statement issued by TÜV Rheinland Energy GmbH dated 12 October 2016

TÜV Rheinland Energy GmbH  
Cologne, 28 March 2017

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl. Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.