

# CONFIRMATION

Notification: 0000051691\_00\_01  
on changes according to EN 15267 regarding certificate 0000051691\_00 dated 19 August 2016

---

**Measuring system:** FWE200DH for dust

**Manufacturer:** SICK Engineering GmbH  
Bergener Ring 27  
01458 Ottendorf-Okrilla  
Germany

---

**German Federal Environmental Agency (UBA)**

**Announcement about the uniform practice in  
monitoring emissions and ambient air  
dated 22 February 2017  
Federal Gazette: BAnz AT 15.03.2017 B6**

**V Notifications to the uniform practice for the continuous monitoring  
of emission and ambient air:**

- 23 Notification as regards Federal Environment Agency notice of 14 July 2016  
(BAnz AT 01.08.2016 B11, chapter I no. 1.2)


The FWE200DH dust measuring system manufactured by SICK Engineering GmbH  
may optionally be equipped with a valve for back purging the sampling line.


Test results are detailed in test report number 936/21232260/A of 10 October 2016  
issued by TÜV Rheinland Energy GmbH.

Statement issued by TÜV Rheinland Energy GmbH dated 10 October 2016

---

TÜV Rheinland Energy GmbH  
Cologne, 28 March 2017

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl. Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.