

CONFIRMATION

Notification: 0000053803_01_01
on changes according to EN 15267 regarding certificate 0000053803_01 dated 08 September 2017

AMS designation: D-R 808 for dust

Manufacturer: DURAG GmbH
Kollastraße 105
22453 Hamburg
Germany

German Federal Environment Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 21 February 2018
Publication: BAnz AT 26.03.2018 B8**

**V. Notifications as regards the uniform practice for the continuous monitoring
of emission and ambient air:**

**25 Notification as regards Federal Environment Agency notice
of 13 July 2017 (BAnz AT 31.07.2017 B12, chapter I number 1.2)**

The current software versions of the D-R 808 measuring system for dust
manufactured by DURAG GmbH are as follows:

D-R 808: 02.10R0002
D-ISC 100: 01.04R0007
D-ESI 100: 01.10R0007


Versions of approved intermediate versions are as follows:


D-ESI 100: 1.2.003

In addition to the existing XPPower DNR240PS24-I power supply, the Phoenix
Contact QUINT4-PS/1AC/24DC/10 power supply may be used for the D-ISC 100.

Statement issued by TÜV Rheinland Energy GmbH dated 8 December 2017

TÜV Rheinland Energy GmbH
Cologne, 4 June 2018


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Phone. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.