

# CONFIRMATION

Notification: 0000053810\_10\_01  
on changes according to EN 15267 regarding certificate 0000053810\_10 dated 05 March 2018

---

**AMS designation:** Set CEM CERT 7MB1957 for CO, NO, SO<sub>2</sub>, CO<sub>2</sub>, NO<sub>2</sub>, NO<sub>x</sub> and O<sub>2</sub>

**Manufacturer:** Siemens AG  
Östliche Rheinbrückenstraße 50  
76187 Karlsruhe  
Germany

---

**German Federal Environment Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 21 February 2018  
Publication: BAnz AT 26.03.2018 B8**

**V. Notifications as regards the uniform practice for the continuous monitoring  
of emission and ambient air:**

**48 Notification as regards Federal Environment Agency (UBA) notice  
of 13 July 2017 (BAnz AT 31.07.2017 B12, chapter I number 3.1)**

The current software versions for the measurement module Ultramat 23 of the Set CEM CERT 7MB1957 measuring system for CO, NO, SO<sub>2</sub>, CO, NO<sub>2</sub>, NO<sub>x</sub> and O<sub>2</sub> manufactured by Siemens AG are as follows:

Ultramat 23-7MB2355	3.00.09
Ultramat 23-7MB2357	3.00.09
Ultramat 23-7MB2358	3.00.09


Software version 3.00.08 has also been approved.


Instead of the analogue KHZ-E-TRV2 doubler, model TTV 2.00 GW may be built into the analyser cabinet of the Set CEM CERT 7MB1957 as an alternative.

Statement issued by TÜV Rheinland Energy GmbH dated 8 December 2017

---

TÜV Rheinland Energy GmbH  
Cologne, 4 June 2018

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Phone. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.