

# CONFIRMATION

Notification: 2640853ts\_01\_02  
on changes according to EN 15267 regarding certificate 2640853ts\_01 dated 20 January 2017

---

**AMS designation:** Advance Optima AO2000 Series for CO, NO, SO<sub>2</sub>, N<sub>2</sub>O, CO<sub>2</sub> and O<sub>2</sub>

**Manufacturer:** ABB Automation GmbH  
Stierstädter Str. 5  
60488 Frankfurt/Main  
Germany

---

**German Federal Environment Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 13 July 2017  
Publication: BANz AT 31.07.2017 B12**

**II. Notifications as regards the uniform practice for the continuous monitoring  
of emission and ambient air:**

**7 Notification as regards Federal Environment Agency (UBA) notices  
of 23 February 2012 (BANz p. 920, chapter I number 4.2) and  
dated 22 February 2017 (BANz AT 15.03.2017 B6, chapter V 17<sup>th</sup> notification)**

The current software versions of the Advance Optima AO2000 measuring CO, NO, SO<sub>2</sub>, N<sub>2</sub>O,  
CO<sub>2</sub> and O<sub>2</sub> manufactured by ABB Automation GmbH are:

AMC board 3.8.0


Syscon: 5.1.12


The Inpotron Type PSU-0261-12-14 power supply may be used as a 24V power supply.

Statement issued by TÜV Rheinland Energy GmbH dated 7 March 2017

---

TÜV Rheinland Energy GmbH  
Cologne, 08 September 2017

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
tre@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Köln

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.