

# CONFIRMATION

Notification: 0000040336\_02\_02  
on changes according to EN 15267 regarding certificate 0000040336\_02 dated 31 May 2022

---

**AMS designation:** Air Pollution Monitor 2 (APM-2) for PM<sub>10</sub> and PM<sub>2,5</sub>

**Manufacturer:** Comde-Derenda GmbH  
Kieler Straße 9  
14532 Stahnsdorf  
Germany

---

**German Federal Environmental Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 19 March 2024  
Publication: BAnz AT 10.05.2024 B7**

**V. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**

**47 Notification as regards Federal Environment Agency (UBA) notices  
of 17 July 2014 (BAnz AT 05.08.2014 B11, chapter III number 2.1) and  
of 21 February 2023 (BAnz AT 20.03.2023 B6, chapter IV notification 63)**

The Air Pollution Monitor 2 (APM-2) measuring system for particulate matter PM<sub>2,5</sub> and PM<sub>10</sub> from Comde-Derenda GmbH can alternatively be equipped with the MPX5100AP absolute pressure sensor from the manufacturer NXP / Freescale Semiconductor.


The current software version is


3.13.001

Statement issued by TÜV Rheinland Energy GmbH dated 23 September 2023

---

TÜV Rheinland Energy & Environment GmbH  
Cologne, 24 May 2024

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.