

CONFIRMATION

Notification: 0000054059_01_05 on changes according to EN 15267 regarding certificate 0000054059_01 dated 5 November 2019

AMS designation:

D-EMS 2020 D-EMS 2020 CS for evaluation system (DAHS)

Manufacturer:

DURAG data systems GmbH

Kollaustr. 105 22453 Hamburg Germany

German Federal Environmental Agency (UBA)

Announcement regarding the uniform practice in monitoring emissions and ambient air dated 19 March 2024
Publication: BAnz AT 10.05.2024 B7

- V. Notifications as regards the uniform practice in monitoring emissions and ambient air:
- 40 Notification as regards Federal Environment Agency (UBA) notices of 28 June 2019 (BAnz AT 22.07.2019 B8, chapter IV number 1.1) and of 9 March 2022 (BAnz AT 11.04.2022 B10, chapter VI notification 44)

The current software version for the D-EMS 2020 emission data evaluation system (ID = 0000054059) from DURAG data systems GmbH is 1.5.2303.19510

The following software versions can also be used: 1.5.2210.18814, 1.5.2210.18886, 1.5.2210.18905, 1.5.2303.19477

In future, the software version number of the D-EMS 2020 will be displayed as follows ${\sf H.N.YYMM.BBBBB}$

Legend: H = Main number, N = Secondary number YY = Year, MM = Month, BBBB = Build number

Furthermore, some errors in the use of import/export functions and when copying raw data to the backup file have been fixed in the D-EMS 2020 emissions data evaluation system. The display and use of the emission data evaluation calculator have also been optimised.

Statement issued by TÜV Rheinland Energy GmbH dated 15 December 2023

TÜV Rheinland Energy & Environment GmbH Cologne, 24 May 2024

i. V. Jol.-Ing. Guido Baum

i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu

Tel. +49 221 806-5200

TÜV Rheinland Energy & Environment GmbH

Am Grauen Stein

51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

05 notification on changes to no. certificate 0000054059_01