

CONFIRMATION

Notification: 0000032298_05_02

on changes according to EN 15267 regarding certificate 0000032298_05 dated 2 March 2023

AMS designation:	D-FL 100 for waste gas velocity
	DURAG GmbH Kollaustr. 105 22453 Hamburg Germany

German Federal Environmental Agency (UBA)

Announcement regarding the uniform practice in monitoring emissions and ambient air dated 19 March 2024 Publication: BAnz AT 10.05.2024 B7

- V. Notifications as regards the uniform practice in monitoring emissions and ambient air:
- 5 Notification as regards Federal Environment Agency (UBA) notices of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter II number 2.4) and of 28 June 2022 (BAnz AT 28.07.2022 B4, chapter III notification 2)

The D-FL 100 waste gas velocity measuring system from DURAG GmbH is now approved for an extended temperature range of -40 °C to +50 °C.

Statement issued by TÜV Rheinland Energy GmbH dated 11 September 2023

TÜV Rheinland Energy & Environment GmbH Cologne, 24 May 2024

i. V. Mpl.-Ing. Guido Baum

etty i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu	TÜV Rheinland Energy & Environment GmbH
TRE@umwelt-tuv.eu	Am Grauen Stein
Tel. +49 221 806-5200	51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

02 notification on changes to no. certificate 0000032298_05