

CONFIRMATION

Notification: 0000081158_01_01 on changes according to EN 15267 regarding certificate 0000081158_01 dated 12 June 2024

AMS designation:

D-R 909 for dust

Manufacturer:

DURAG GmbH Kollaustr. 105 22453 Hamburg Germany

German Federal Environmental Agency (UBA)

Announcement regarding the uniform practice in monitoring emissions and ambient air dated 31 August 2024
Publication: BAnz AT 31.10.2024 B9

- IV. Notifications as regards the uniform practice in monitoring emissions and ambient air:
- 17 Notification as regards Federal Environment Agency (UBA) notice of 19 March 2024 (BAnz AT 10.05.2024 B7, chapter I number 2.2)

The D-R 909 measuring system for dust from DURAG GmbH is equipped with three additional relays (4039288 relay 24VDC 50mA 2 PDT) to adapt the compressor control in order to cover a wider range of sample gas pressures (especially negative pressure). In this context, the software has also been adapted accordingly.

The new software version is:

01.05R0596.

Statement issued by TÜV Rheinland Energy & Environment GmbH dated 10 May 2024

TÜV Rheinland Energy & Environment GmbH Cologne, 12 November 2024

i. V. Jol.-Ing. Guido Baum

i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu

TRE@umwelt-tuv.eu Tel. +49 221 806-5200 TÜV Rheinland Energy & Environment GmbH

Am Grauen Stein 51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.