

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000040328_03

Certified AMS: AC32M for NO, NO₂ and NO_x

Manufacturer: ENVEA
111, Boulevard Robespierre
78304 Poissy Cedex
France

Test Institute: TÜV Rheinland Energy & Environment GmbH

**This is to certify that the AMS has been tested
and found to comply with the standards
VDI 4202-1 (2002), VDI 4203-3 (2004), EN 14211 (2012)
as well as EN 15267-1 (2009) and EN 15267-2 (2023).**

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 11 pages).

The present certificate replaces certificate 0000040328_02 dated 1 July 2020.



Suitability Tested
Complying with
2008/50/EC
EN 15267
Regular
Surveillance

www.tuv.com
ID 0000040328


Publication in the German Federal Gazette
(BAnz) of 20 April 2007

German Environment Agency


Dessau, 27 June 2025

This certificate will expire on:
30 June 2030

TÜV Rheinland Energy &
Environment GmbH
Cologne, 26 June 2025



Dr. Marcel Langner
Head of Section II 4



ppa. Dr. Peter Wilbring

www.umwelt-tuv.eu
qal1-info@tuv.com
Tel. + 49 221 806-5200

TÜV Rheinland Energy & Environment GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

Test report:	936/21205818/A dated 8 December 2006 and Addendum 936/21221709/A dated 28 September 2013
Initial certification:	1 April 2014
Expiry date:	30 June 2030
Certificate:	Renewal (of previous certificate 0000040328_02 of 1 July 2020 valid until 30 June 2025)
Publication:	BAnz. 20 April 2007, No. 75, p. 4139, chapter III No. 4.1 and Banz AT 01.04.2014 B12, chapter VI notification 18

Approved application

The tested AMS is suitable for continuous immission measurement of NO, NO₂ and NO_x in stationary use.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three month field test.

The AMS is approved for an ambient temperature range of 0 °C to 30 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the measured values relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the intended use.

Basis of the certification

This certification is based on:

- Test report 936/21205818/A dated 8 December 2006 of TÜV Rheinland Immissionsschutz und Energiesysteme GmbH and Addendum 936/21221709/A dated 28 September 2013 issued by TÜV Rheinland Energie und Umwelt GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz. 20 April 2007, No. 75, p. 4139, chapter III No. 4.1, Announcement by UBA dated 12 April 2007:

AMS designation:

AC32M for NO, NO₂ and NO_x

Manufacturer:

Environnement S.A., Poissy Cedex, France and
Ansyco GmbH Karlsruhe, Germany

Field of application:

For continuous monitoring of NO, NO₂ and NO_x in ambient air (stationary operation)

Measuring ranges during the performance test:

NO ₂	0 – 400	µg/m ³
NO ₂	0 – 500	µg/m ³
NO	0 – 1200	µg/m ³

Software version:

V2.45

Test institute:

TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Cologne
TÜV Rheinland Group
Report No.: 936/21205818/A dated 8 December 2006

Publication in the German Federal Gazette: BAnz AT 01.04.2014 B12, Chap. VI notification 18, Announcement by UBA dated 27 February 2014:

18 Notification of announcement by the German Federal Environment Agency of 12 April 2007 (BAnz. p. 4139, Chapter III Number 4.1)

The measuring system AC32M for NO, NO₂ and NO_x manufactured by Environnement fulfils the requirements of DIN EN 14211 (November 2012). Furthermore, the manufacturing process and quality management system of the measuring system AC32M for NO, NO₂ and NO_x fulfill the requirements of EN 15267.

The test report of the performance test with report number 936/21205818/A and an addendum as an integral part of to the test report with report number 936/21221709/A can be viewed on the internet at www.qal1.de.

Statement by TÜV Rheinland Energie und Umwelt GmbH dated 28 September 2013

Publication in the German Federal Gazette: BAnz AT 26.08.2015 B4, Chap. V notification 51,
Announcement by UBA dated 22 July 2015:

**51 Notification as regards Federal Environment Agency (UBA) notices
of 12 April 2007 (BAnz. p. 4139, chapter III number 4.1) and
of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter VI notification 18)**

The current software version for the AC32M measuring system for NO, NO₂ and NO_x,
manufactured by Environnement S.A., is:

v1.02 (Calculation process)

v3.6.b (Display process)

For additional means of communication, the AMS was equipped with a USB and a
TCP/IP interface.

The ozone generator was optimised and now carries the designation B01-5005-1.

The pressure sensors in the measurement chamber and within the sample gas
stream have been replaced in each case by one sensor of the type C06-C5-0291-A.

The high voltage supply was replaced by one of the type PS1800N/12F.

Statement by TÜV Rheinland Energie und Umwelt GmbH of 14 March 2015

Publication in the German Federal Gazette: BAnz AT 26.03.2019 B7, Chap. IV
notification 21, Announcement by UBA dated 27 February 2019:

**21 Notification as regards Federal Environment Agency notices
of 12 April 2007 (BAnz. p. 4139, chapter III number 4.1) and
of 22 July 2015 (BAnz AT 26.08.2015 B4, chapter V notification 51)**

The current software version of the AC32M measuring system
for NO, NO₂ and NO_x manufactured by Environnement S.A. is:

v1.02 (calculation process)

v3.6.g (display process)

Statement issued by TÜV Rheinland Energy GmbH dated 27 September 2018

Publication in the German Federal Gazette: BAnz AT 24.03.2020 B7, Chap. IV
notification 25, Announcement by UBA dated 24 February 2020:

**25 Notification as regards Federal Environment Agency (UBA) notices
of 12 April 2007 (BAnz. p. 4139, chapter III number 4.1) and
of 27 February 2019 (BAnz AT 26.03.2019 B7, chapter IV notification 21)**


Environnement S.A., Poissy, France have changed their company name to ENVEA.
The latest software version of the AC32M measuring system for NO, NO₂ und NO_x
manufactured by ENVEA is
v1.02 (calculation process)
v3.8.b (display process)

Statement issued by TÜV Rheinland Energy GmbH dated 1 October 2019

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The AC32M analyser measures nitrogen oxide (NO) and nitrogen dioxide (NO₂) in ambient air. The measuring principle is based on the light emission from the chemical reaction between NO and ozone in the reaction chamber, also referred to as chemiluminescence. The chemiluminescence represents the oxidation of NO molecules by ozone molecules to NO₂* molecules.



The return of the excited NO₂* molecules to a basic electronic condition is performed using light radiation in a spectrum of 600 to 1200 nanometres:



This energy can be lost (quenching) in the sample due to a collision with certain molecules, in particular H₂O and CO₂. A reduction of the pressure in the reaction chamber to approx. 200 mbar and the drying of the sample using a Perma Pure dryer reduces the likelihood of a collision which means a greater light yield and thus detection limit can be achieved.

The ozone required is generated in an internal ozone generator using stationary electrical charging in a cylindrical capacitor.

The reaction chamber is separated from the sensor by an optical red filter which only allows light beams with a wavelength of over 610 nanometres to pass through and thus suppresses the inferences caused by the hydrocarbons.

Radiation measurement is performed using a photomultiplier (PM). The electrical signal it provides is magnified and digitised for processing by the microprocessor.

In order to be measured by chemiluminescence the NO₂ must first be converted to NO. A hot molybdenum converter is used to perform this reduction according to the following reaction equation:



The sample is sucked in by a vacuum pump at the entry to the system which provides the vacuum in the reaction chamber and the Perma Pure dryer blowback.

The measurement comprises 3 cycles:

- **Reference cycle:** The sample is fed into a pre-reaction chamber (tube section) where it is mixed with ozone. The NO in the sample is oxidised into NO₂ before it flows into the reaction chamber. The signal measured by the PM without chemiluminescence can be considered a measurement with zero air and serves as a reference signal or zero signal.
- **NO cycle:** The sample is fed directly into the measuring chamber in which the NO molecules are oxidised with ozone. The signal measured by the PM is proportional to the number of NO molecules present in the sample.
- **NO_x cycle:** The sample is fed through the NO₂ converter and then mixed with ozone in the reaction chamber. The signal measured by the PM is proportional to the total number of NO and NO₂ molecules in the sample, the latter originating from the reduction of NO₂. The sum of NO + NO₂ is referred to as NO_x.

This measuring principle corresponds to the standard reference method as described in EN 14211.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy & Environment GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy & Environment GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy & Environment GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: gal1.de.

History of documents

Certification of AC32M is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Basic test

Test report: 936/21205818/A dated 8 December 2006
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH
Publication: BAnz. 20 April 2007, No. 75, p. 4139, chapter III number 4.1
UBA announcement dated 12 April 2007

Initial certification according to EN 15267 based on notification

Certificate No. 0000040328_00: 29 April 2014
Expiry date of the certificate: 31 March 2019
Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 28 September 2013
Test report: 936/21205818/A dated 8 December 2006 and Addendum 936/21221709/A dated 28 September 2013
Publication: BAnz AT 01.04.2014 B12, chapter VI notification 18
UBA announcement dated 27 February 2014

Notifications

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 14 March 2015
Publication: BAnz AT 26.08.2015 B4, chapter V notification 51
UBA announcement dated 22 July 2015
(Soft- and hardware changes)

Renewal of certificates

Certificate No. 0000040328_01: 1 April 2019
Expiry date of the certificate: 30 June 2020

Notifications

Statement issued by TÜV Rheinland Energy GmbH dated 27 September 2018
Publication: BAnz AT 26.03.2019 B7, chapter IV notification 21
UBA announcement dated 27 February 2019
(Software changes)

Statement issued by TÜV Rheinland Energy GmbH dated 1 October 2019
Publication: BAnz AT 24.03.2020 B7, chapter IV notification 25
UBA announcement dated 24 February 2020
(Software changes and new producer name formerly Environnement S.A.)

Renewal of certificates

Certificate No. 0000040328_02: 1 July 2020
Expiry date of the certificate: 30 June 2025

Renewal of certificates

Certificate No. 0000040328_03: 27 June 2025
Expiry date of the certificate: 30 June 2030

Expanded uncertainty laboratory, system 1

Measuring device:		Environnement AC32M		Serial-No.:		Gerät 1	
Measured component:		NO ₂		1h-limit value:		104.6	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty	nmol/mol	
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.610	U _{r,z}	0.10	0.0101	
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	2.260	U _{r,1h}	0.08	0.0059	
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	0.300	U _{l,1h}	0.18	0.0328	
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 8.0 nmol/mol/kPa	0.250	U _{gp}	0.57	0.3205	
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 3.0 nmol/mol/K	-0.060	U _{gt}	-0.14	0.0205	
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 3.0 nmol/mol/K	0.200	U _{st}	0.52	0.2679	
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.086	U _v	0.26	0.0684	
8a	Interferent H ₂ O with 21 mmol/mol	≤ 10 nmol/mol (Zero) ≤ 10 nmol/mol (Span)	2.300 1.700	U _{H2O}	1.62	2.6327	
8b	Interferent CO ₂ with 500 µmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	0.700 2.000	U _{int,pos} or	0.86	0.7313	
8c	Interferent NH ₃ mit 200 nmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	0.300 1.300	U _{int,neg}			
9	Averaging effect	≤ 7.0% of measured value	1.900	U _{av}	1.15	1.3166	
18	Difference sample/calibration port	≤ 1.0%	0.000	U _{sc}	0.00	0.0000	
21	Converter efficiency	≥ 98	98.40	U _{ec}	1.67	2.8009	
23	Uncertainty of test gas	≤ 3.0%	2.000	U _{cg}	1.05	1.0941	
Combined standard uncertainty				u _c		3.0525	
Expanded uncertainty				U		6.1051	
Relative expanded uncertainty				W		5.84	
Maximum allowed expanded uncertainty				W _{req}		15	

Expanded uncertainty laboratory, system 2

Measuring device:		Serial-No.:		Gerät 2	
Measured component:		1h-limit value:		104.6	
NO ₂				nmol/mol	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.640	U _{r,z}	0.0121
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	2.700	U _{r,1h}	0.0092
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	0.300	U _{l,1h}	0.0328
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 8.0 nmol/mol/kPa	0.200	U _{gp}	0.2051
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 3.0 nmol/mol/K	0.060	U _{gt}	0.0205
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 3.0 nmol/mol/K	0.200	U _{st}	0.2679
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	-0.029	U _v	0.0078
8a	Interferent H ₂ O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	0.000	U _{H2O}	1.4546
8b	Interferent CO ₂ with 500 µmol/mol	≤ 10 nmol/mol (Span)	0.000	U _{Int,pos} or U _{Int,neg}	0.8758
8c	Interferent NH ₃ mit 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	0.300		
		≤ 5.0 nmol/mol (Span)	0.000		
		≤ 5.0 nmol/mol (Zero)	1.300		
9	Averaging effect	≤ 7.0% of measured value	1.700	U _{av}	0.0584
18	Difference sample/calibration port	≤ 1.0%	0.000	U _{Δsc}	0.0000
21	Converter efficiency	≥ 98	98.80	U _{EC}	1.5755
23	Uncertainty of test gas	≤ 3.0%	2.000	U _{cg}	1.0941
Combined standard uncertainty			U _c	2.3738	nmol/mol
Expanded uncertainty			U	4.7477	nmol/mol
Relative expanded uncertainty			W	4.54	%
Maximum allowed expanded uncertainty			W _{req}	15	%

Combined uncertainty, laboratory and field, system 1

Measuring device:		Environnement AC32M		Serial-No.:		Gerät 1	
Measured component:		NO ₂		1h-limit value:		104.6	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty	nmol/mol	
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.610	u _{r,z}	0.10	0.0101	
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	2.260	u _{r,1h}	not considered, as $\sqrt{2} \cdot u_{r,1h} = 0.1 < u_{r,f}$	-	
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	0.300	u _{l,1h}	0.18	0.0328	
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 8.0 nmol/mol/kPa	0.250	u _{pp}	0.57	0.3205	
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 3.0 nmol/mol/K	-0.060	u _{gt}	-0.14	0.0205	
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 3.0 nmol/mol/K	0.200	u _{st}	0.52	0.2679	
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.086	u _y	0.26	0.0684	
8a	Interferent H ₂ O with 21 mmol/mol	≤ 10 nmol/mol (Zero) ≤ 10 nmol/mol (Span)	2.300 1.700	u _{H2O}	1.62	2.6327	
8b	Interferent CO ₂ with 500 µmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	0.700 2.000	u _{CO2,pos} or	0.86	0.7313	
8c	Interferent NH ₃ mit 200 nmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	0.300 1.300	u _{NH3,neg}			
9	Averaging effect	≤ 7.0% of measured value	1.900	u _{av}	1.15	1.3166	
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	4.890	u _{r,f}	5.11	26.1626	
11	Long term drift at zero level	≤ 5.0 nmol/mol	-0.590	u _{d,l,z}	-0.34	0.1160	
12	Long term drift at span level	≤ 5.0% of max. of certification range	0.790	u _{d,l,1h}	0.48	0.2276	
18	Difference sample/calibration port	≤ 1.0%	0.000	u _{asc}	0.00	0.0000	
21	Converter efficiency	≥ 98	98.400	u _{ec}	1.67	2.8009	
23	Uncertainty of test gas	≤ 3.0%	2.000	u _{cg}	1.05	1.0941	
Combined standard uncertainty				u _c		5.9843	
Expanded uncertainty				U		11.9687	
Relative expanded uncertainty				W		11.44	
Maximum allowed expanded uncertainty				W _{req}		15	

Combined uncertainty, laboratory and field, system 2

Measuring device:		Environnement AC32M		Serial-No.:		Gerät 2	
Measured component:		NO ₂		1h-limit value:		104.6	
						nmol/mol	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty		
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.640	u _{r,z}	0.11		
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	2.700	u _{r,h}	not considered, as $\sqrt{2} \cdot u_{r,h} = 0.13 < u_{r,f}$	-	
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	0.300	u _{l,h}	0.18	0.0328	
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 8.0 nmol/mol/kPa	0.200	u _{pp}	0.45	0.2051	
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 3.0 nmol/mol/K	0.060	u _{gt}	0.14	0.0205	
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 3.0 nmol/mol/K	0.200	u _{st}	0.52	0.2679	
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	-0.029	u _v	-0.09	0.0078	
8a	Interferent H ₂ O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	1.700	u _{H2O}	1.21	1.4546	
8b	Interferent CO ₂ with 500 µmol/mol	≤ 10 nmol/mol (Span)	1.300	u _{int,pos} or u _{int,neg}	0.94	0.8758	
8c	Interferent NH ₃ mit 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	0.300				
9	Averaging effect	≤ 5.0 nmol/mol (Span)	1.300				
10	Reproducibility standard deviation under field conditions	≤ 7.0% of measured value	0.400	u _{av}	0.24	0.0584	
11	Long term drift at zero level	≤ 5.0% of average over 3 months	4.890	u _{rl}	5.11	26.1626	
12	Long term drift at span level	≤ 5.0 nmol/mol	0.780	u _{d,l,z}	0.45	0.2028	
18	Difference sample/calibration port	≤ 5.0% of max. of certification range	0.660	u _{d,l,h}	0.40	0.1589	
21	Converter efficiency	≤ 1.0%	0.000	u _{asc}	0.00	0.0000	
23	Uncertainty of test gas	≥ 98	98.800	u _{ec}	1.26	1.5755	
				u _{cg}	1.05	1.0941	
				Combined standard uncertainty		5.6693	
				Expanded uncertainty		11.3386	
				Relative expanded uncertainty		10.84	
				Maximum allowed expanded uncertainty		15	