

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000040330_02

AMS designation: CO12M for CO

Manufacturer: ENVEA
111 Boulevard Robespierre
78304 Poissy Cedex
France

Test Laboratory: TÜV Rheinland Energy GmbH

**This is to certify that the AMS has been tested
and found to comply with the standards:
VDI 4202-1 (2002), VDI 4203-3 (2004), EN 14626 (2012),
EN 15267-1 (2009) and EN 15267-2 (2009).**

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 11 pages).
The present certificate replaces certificate 0000040330_01 of 01 April 2019.



Suitability Tested
Equivalent to
2008/50/EC
EN 15267
Regular Surveillance
www.tuv.com
ID 0000040330

Publication in the German Federal Gazette
(BAnz) of 01 April 2014

This certificate will expire on:
30 June 2025

German Federal Environment Agency
Dessau, 01 July 2020

TÜV Rheinland Energy GmbH
Cologne, 30 June 2020



Dr. Marcel Langner
Head of Section II 4.1



ppa. Dr. Peter Wilbring

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Phone: + 49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to certificate D-PL-11120-02-00.

Test Report:	936/21206773/B dated 29 February 2008
Initial certification:	01 April 2014
Expiry date:	30 June 2025
Certificate:	Renewal (of previous certificate 0000040330_01 dated 01 April 2019 valid until 30 June 2020)
Publication:	BAnz AT 01.04.2014 B12, chapter VI notification 20

Approved application

The certified AMS is suitable for continuous monitoring of carbon monoxide in ambient air.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three-months field test.

The AMS is approved for an ambient temperature range of 0 °C to +30 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for monitoring the AMS readings relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the intended purpose.

Basis of the certification

This certification is based on:

- Test report no. 936/21206773/B dated 29 February 2008 issued by TÜV Rheinland Immissionschutz und Energiesysteme GmbH and Addendum 936/21221709/D dated 28 September 2013 issued by TÜV Rheinland Energie und Umwelt GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz. 03 September 2008, no. 133, p. 3243, chapter III number 1.1, UBA announcement dated 12 August 2008:

AMS designation:

CO12M for CO

Manufacturer:

Environnement S.A., Poissy Cedex, France

Distribution in Germany:

Ansyco GmbH, Karlsruhe

Field of application:

For the continuous monitoring of carbon monoxide in ambient air

Measuring range during performance testing:

CO 0 to 60 mg/m³

0 to 100 mg/m³

Software version:

V1.26

Test Laboratory:

TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Cologne

TÜV Rheinland Group

Report no.: 936/21206773/B dated 29 February 2008

Publication in the German Federal Gazette: BAnz AT 01.04.2014 B12, chapter VI notification 20, UBA announcement dated 27 February 2014:

20 Notification as regards Federal Environment Agency (UBA) notices of 12 August 2008 (BAnz. p. 3243, chapter III number 1.1)

The Model CO12M air quality monitor for CO manufactured by Environnement complies with the requirements of EN 14626 (August 2012 version). Furthermore, the manufacturing process and the quality management for the Model CO12M measuring system for CO meet the requirements of EN 15267.

The test report on performance testing No. 936/21206773/B and the addendum no. 936/21221709/D as an integral part of this report are available online at www.qal1.de.

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 28 September 2013

Publication in the German Federal Gazette: BAnz AT 26.08.2015 B4, chapter V notification 48, UBA announcement dated 22 July 2015:

48 Notification as regards Federal Environment Agency (UBA) notices of 12 August 2008 (BAnz. p. 3243, chapter III number 1.1) and of 27 February 2014 (BAnz AT 01.04.2014 B12, chapter VI 20th notification)

The latest software version of the CO12M measuring system for CO manufactured by Environnement S.A. is:

v1.0.d (calculation process)
v3.6.c (display process)

To extend the means of communication, the measuring system will be equipped with a USB port and a TCP/IP interface.
The Deltaline 26BC-6A-107.101 filter wheel motor replaces the previous Maxon DC Type A-max 22.

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 16 March 2015

Publication in the German Federal Gazette: BAnz AT 26.03.2019 B7, chapter IV notification 25, UBA announcement dated 27 February 2019:

25 Notification as regards Federal Environment Agency (UBA) notices of 12 August 2008 (BAnz. p. 3243, chapter III number 1.1) and of 22 July 2015 (BAnz AT 26.08.2015 B4, chapter V 48th notification)

The latest software version of the CO12M measuring system for CO manufactured by Environnement S.A. is:

v1.0.d (calculation process)
v3.6.f (display process)

Statement issued by TÜV Rheinland Energy GmbH dated 27 September 2018

Publication in the German Federal Gazette: BAnz AT 24.03.2020 B7, chapter IV notification 30, UBA announcement dated 24 February 2020:

30 Notification as regards Federal Environment Agency (UBA) notices of 12 August 2008 (BAnz. p. 3243, chapter III number 1.1) and of 27 February 2019 (BAnz AT 26.03.2019 B7, chapter IV 25th notification)

Environnement S.A., Poissy, France have changed their company name to ENVEA.

The latest software version of the CO12M measuring system for CO manufactured by ENVEA is:
v1.0.d (calculation process)
v3.8.a (display process)

Statement issued by TÜV Rheinland Energy GmbH dated 1 October 2019

Certified product

This certification applies to automated measurement systems conforming to the following description:

The CO12M analyser measures carbon monoxide (CO) in ambient air. The measuring principle is based on infra-red absorption according to the Beer-Lambert law. The absorption spectrum of carbon monoxide has a maximum wavelength of 4.67 μm , which corresponds to the spectrum selected by the optical filter.

Due to the fact that the absorption spectrum is not continuous, the optical filter is connected to a gas filter correlation wheel which enables highly selective measurement of the gas to be analysed by eliminating interferences caused by gases with similar absorption spectrums to those of CO.

This measuring principle corresponds to the standard reference method as described in EN 14626.

General remarks

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacturing process for the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. Upon revocation of the publication the certificate loses its validity. After the expiration of the certificate and on request of TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must no longer be used.

The relevant version of this certificate and its expiration date are also accessible on the internet at qal1.de.

Document history

Certification of the CO12M measuring system is based on the documents listed below and the regular, continuous surveillance of the manufacturer's quality management system:

Basic testing

Test Report: 936/21206773/B dated 29 February 2008
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH
Publication: BAnz. 03 September 2008, no. 133, p. 3243, chapter III number 1.1
UBA announcement dated 12 August 2008

Initial certification according to EN 15267

Certificate no. 0000040330: 29 April 2014
Expiry date of the certificate: 31 March 2019
Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 28 September 2013
Test Report: 936/21221709/D dated 28 September 2013
Publication: BAnz AT 01.04.2014 B12, chapter VI notification 20
UBA announcement dated 27 February 2014

Notifications in accordance with EN 15267

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 16 March 2015
Publication: BAnz AT 26.08.2015 B4, chapter V notification 48
UBA announcement dated 22 July 2015
(Design and software changes)

Renewal of the certificate

Certificate no. 0000040330_01: 01 April 2019
Expiry date of the certificate: 30 June 2020

Notifications in accordance with EN 15267

Statement issued by TÜV Rheinland Energy GmbH dated 27 September 2018
Publication: BAnz AT 26.03.2019 B7, chapter IV notification 25
UBA announcement dated 27 February 2019
(software updates)

Statement issued by TÜV Rheinland Energy GmbH dated 1 October 2019
Publication: BAnz AT 24.03.2020 B7, chapter IV notification 30
UBA announcement dated 24 February 2020
(New company name, software changes)

Renewal of the certificate

Certificate no. 0000040330_02: 01 July 2020
Expiry date of the certificate: 30 June 2025

Expanded uncertainty from the results obtained in the laboratory tests for analyser 1

Measuring device:		Serial-No.:		Gerät 1	
Environment CO12M		8h-limit value:		8.62	
CO					
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at zero	≤ 0.3 µmol/mol	0.000	u _{r,z}	0.0000
2	Repeatability standard deviation at 8h-limit value	≤ 0.4 µmol/mol	0.200	u _r	0.0010
3	"lack of fit" at 8h-limit value	≤ 4.0% of measured value	1.100	u _f	0.0030
4	Sensitivity coefficient of sample gas pressure at 8h-limit value	≤ 0.7 µmol/mol/kPa	0.010	u _{sp}	0.0005
5	Sensitivity coefficient of sample gas temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.015	u _{gt}	0.0010
6	Sensitivity coefficient of surrounding temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.050	u _{st}	0.0114
7	Sensitivity coefficient of electrical voltage at 8h-limit value	≤ 0.3 µmol/mol/V	0.000	u _v	0.0000
8a	Interferent H ₂ O with 21 mmol/mol	≤ 1.0 µmol/mol (Zero)	0.060	u _{H2O}	0.0017
		≤ 1.0 µmol/mol (Span)	-0.040		
8b	Interferent CO ₂ with 500 µmol/mol	≤ 0.5 µmol/mol (Zero)	-0.090	u _{int,pos}	
		≤ 0.5 µmol/mol (Span)	-0.100		
8c	Interferent NO with 1 µmol/mol	≤ 0.5 µmol/mol (Zero)	0.020		
		≤ 0.5 µmol/mol (Span)	0.130	or	0.0229
8d	Interferent N ₂ O with 50 nmol/mol	≤ 0.5 µmol/mol (Zero)	0.060	u _{int,neg}	
		≤ 0.5 µmol/mol (Span)	0.140		
9	Averaging effect	≤ 7.0% of measured value	-3.900	u _{av}	0.0377
18	Difference sample/calibration port	≤ 1.0%	-0.120	u _{asc}	0.0001
21	Uncertainty of test gas	≤ 3.0%	2.000	u _{cg}	0.0074
				Combined standard uncertainty	
				u _c	0.2944
				Expanded uncertainty	
				U	0.5889
				Relative expanded uncertainty	
				W	6.83
				Maximum allowed expanded uncertainty	
				W _{req}	15

Expanded uncertainty from the results obtained in the laboratory tests for analyser 2

Measuring device:		Environment CO12M		Serial-No.:		Gerät 2	
Measured component:		CO		8h-limit value:		8.62	
						µmol/mol	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty		
1	Repeatability standard deviation at zero	≤ 0.3 µmol/mol	0.200	U _{r,z}	0.04	0.0014	
2	Repeatability standard deviation at 8h-limit value	≤ 0.4 µmol/mol	0.100	U _r	0.02	0.0003	
3	"lack of fit" at 8h-limit value	≤ 4.0% of measured value	0.600	U _l	0.03	0.0009	
4	Sensitivity coefficient of sample gas pressure at 8h-limit value	≤ 0.7 µmol/mol/kPa	0.010	U _{gp}	0.02	0.0005	
5	Sensitivity coefficient of sample gas temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.010	U _{gt}	0.02	0.0005	
6	Sensitivity coefficient of surrounding temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.020	U _{st}	0.04	0.0018	
7	Sensitivity coefficient of electrical voltage at 8h-limit value	≤ 0.3 µmol/mol/V	0.000	U _v	0.00	0.0000	
8a	Interferent H ₂ O with 21 mmol/mol	≤ 1.0 µmol/mol (Zero) ≤ 1.0 µmol/mol (Span)	0.030 -0.070	U _{H2O}	0.02	0.0004	
8b	Interferent CO ₂ with 500 µmol/mol	≤ 0.5 µmol/mol (Zero) ≤ 0.5 µmol/mol (Span)	0.100 -0.140	U _{hi,pos}			
8c	Interferent NO with 1 µmol/mol	≤ 0.5 µmol/mol (Zero) ≤ 0.5 µmol/mol (Span)	-0.060 0.100	or	0.07	0.0056	
8d	Interferent N ₂ O with 50 nmol/mol	≤ 0.5 µmol/mol (Zero) ≤ 0.5 µmol/mol (Span)	-0.040 0.040	U _{hi,neg}			
9	Averaging effect	≤ 7.0% of measured value	-1.300	U _{av}	-0.06	0.0042	
18	Difference sample/calibration port	≤ 1.0%	-0.050	U _{sc}	0.00	0.0000	
21	Uncertainty of test gas	≤ 3.0%	2.000	U _{cg}	0.09	0.0074	
				Combined standard uncertainty		u _c	0.1514
				Expanded uncertainty		U	0.3027
				Relative expanded uncertainty		W	3.51
				Maximum allowed expanded uncertainty		W _{req}	15

Expanded uncertainty from the results obtained in the laboratory and field tests for analyser 1

Measuring device: Environment CO12M		Serial.No.: Gerät 1		µmol/mol	
Measured component: CO		8h-limit value:		8.62	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at zero	≤ 0.3 µmol/mol	0.000	U _{r,z}	0.0000
2	Repeatability standard deviation at 8h-limit value	≤ 0.4 µmol/mol	0.200	U _r	-
3	"lack of fit" at 8h-limit value	≤ 4.0% of measured value	1.100	U _i	0.0030
4	Sensitivity coefficient of sample gas pressure at 8h-limit value	≤ 0.7 µmol/mol/kPa	0.010	U _{sp}	0.0005
5	Sensitivity coefficient of sample gas temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.015	U _{gt}	0.0010
6	Sensitivity coefficient of surrounding temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.050	U _{st}	0.0114
7	Sensitivity coefficient of electrical voltage at 8h-limit value	≤ 0.3 µmol/mol/V	0.000	U _v	0.0000
8a	Interferent H ₂ O with 21 mmol/mol	≤ 1.0 µmol/mol (Zero) ≤ 1.0 µmol/mol (Span)	-0.040 0.060	U _{H2O}	0.0017
8b	Interferent CO ₂ with 500 µmol/mol	≤ 0.5 µmol/mol (Zero) ≤ 0.5 µmol/mol (Span)	-0.090 -0.100	U _{int, pos}	
8c	Interferent NO with 1 µmol/mol	≤ 0.5 µmol/mol (Zero) ≤ 0.5 µmol/mol (Span)	0.020 0.130	or	0.0229
8d	Interferent N ₂ O with 50 mmol/mol	≤ 0.5 µmol/mol (Zero) ≤ 0.5 µmol/mol (Span)	0.060 0.140	U _{int, neg}	
9	Averaging effect	≤ 7.0% of measured value	-3.900	U _{av}	0.0377
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	3.270	U _{r,f}	0.0795
11	Long term drift at zero level	≤ 0.5 µmol/mol	0.220	U _{d,i,z}	0.0161
12	Long term drift at span level	≤ 5.0% of max. of certification range	0.940	U _{d,i,sh}	0.0022
18	Difference sample/calibration port	≤ 1.0%	-0.120	U _{asc}	0.0001
21	Uncertainty of test gas	≤ 3.0%	2.000	U _{cg}	0.0074
Combined standard uncertainty				U _c	0.4283
Expanded uncertainty				U	0.8566
Relative expanded uncertainty				W	9.94
Maximum allowed expanded uncertainty				W _{req}	15

Expanded uncertainty from the results obtained in the laboratory and field tests for analyser 2

Measuring device: Umwelt CO12M		Serial-No.: Gerät 2		µmol/mol	
Measured component: CO		8h-limit value:		8.62	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at zero	≤ 0.3 µmol/mol	0.200	u _{r,z}	0.04
2	Repeatability standard deviation at 8h-limit value	≤ 0.4 µmol/mol	0.100	u _r	not considered, as u _r = 0.01 < u _{r,f}
3	"lack of fit" at 8h-limit value	≤ 4.0% of measured value	0.600	u _i	0.0009
4	Sensitivity coefficient of sample gas pressure at 8h-limit value	≤ 0.7 µmol/mol/kPa	0.010	u _{gp}	0.0005
5	Sensitivity coefficient of sample gas temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.010	u _{gt}	0.0005
6	Sensitivity coefficient of surrounding temperature at 8h-limit value	≤ 0.3 µmol/mol/K	0.020	u _{st}	0.0018
7	Sensitivity coefficient of electrical voltage at 8h-limit value	≤ 0.3 µmol/mol/V	0.000	u _v	0.0000
8a	Interferent H ₂ O with 21 mmol/mol	≤ 1.0 µmol/mol (Zero)	-0.070	u _{H2O}	0.0004
		≤ 1.0 µmol/mol (Span)	0.030		
8b	Interferent CO ₂ with 500 µmol/mol	≤ 0.5 µmol/mol (Zero)	0.100	u _{int,pos}	
		≤ 0.5 µmol/mol (Span)	-0.140		
8c	Interferent NO with 1 µmol/mol	≤ 0.5 µmol/mol (Zero)	-0.060		
		≤ 0.5 µmol/mol (Span)	0.100	or	0.0056
8d	Interferent N ₂ O with 50 nmol/mol	≤ 0.5 µmol/mol (Zero)	-0.040		
		≤ 0.5 µmol/mol (Span)	0.040	u _{int,neg}	
9	Averaging effect	≤ 7.0% of measured value	-1.300	u _{av}	0.0042
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	3.270	u _{r,f}	0.0795
11	Long term drift at zero level	≤ 0.5 µmol/mol	0.380	u _{d,l,z}	0.0481
12	Long term drift at span level	≤ 5.0% of max. of certification range	1.380	u _{d,l,sh}	0.0047
18	Difference sample/calibration port	≤ 1.0%	-0.050	u _{ssc}	0.0000
21	Uncertainty of test gas	≤ 3.0%	2.000	u _{cg}	0.0074
Combined standard uncertainty			u _c		µmol/mol
Expanded uncertainty			U		µmol/mol
Relative expanded uncertainty			W		%
Maximum allowed expanded uncertainty			W _{req}		%