

# CERTIFICATE

## of Product Conformity (QAL1)

Certificate No.: 0000040331\_01

---

**Certified AMS:** O342M for O<sub>3</sub>

**Manufacturer:** Environnement S.A.  
111 Boulevard Robespierre  
78304 Poissy Cedex  
France

---

**Test Institute:** TÜV Rheinland Energy GmbH

**This is to certify that the AMS has been tested  
and found to comply with:**

**VDI 4202-1: 2002, VDI 4203-2: 2004, EN 14625: 2012,  
EN 15267-1: 2009, EN 15267-2: 2009**

Certification is awarded in respect of the conditions stated in this certificate  
(see also the following pages).

The present certificate replaces certificate 0000040331 of 29 April 2014.




Suitability Tested  
Complying with  
2008/50/EC  
EN 15267  
Regular  
Surveillance

www.tuv.com  
ID 0000040331

Publication in the German Federal Gazette  
(BAnz.) of 20 April 2007


German Federal Environment Agency  
Dessau, 1 April 2019



Dr. Marcel Langner  
Head of Section II 4.1

This certificate will expire on:  
30 June 2020

TÜV Rheinland Energy GmbH  
Cologne, 31 March 2019



ppa. Dr. Peter Wilbring

[www.umwelt-tuv.de](http://www.umwelt-tuv.de) / [www.enviro-tuv.eu](http://www.enviro-tuv.eu)  
tre@umwelt-tuv.de  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Accreditation according to EN ISO/IEC 17025:2018 and certified according to ISO 9001:2015.

**Certificate:**  
0000040331\_01 / 1 April 2019

**Test report:** 936/21205818/B of 08 December 2006  
Addendum 936/21221709/C of 28 September 2013

**Initial certification:** 01 April 2014

**Date of expiry:** 30 June 2020

**Publication:** BAnz AT 01 April 2014 B12, chapter VI, notification 21

**Approved application**

The certified AMS is suitable for continuous monitoring of ozone in ambient air.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three-month field test.

The AMS is approved for a temperature range of 0 °C to +30 °C.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for ambient air applications at which it will be installed.

**Basis of the certification**

This certification is based on:

- test report 936/21205818/B of 08 December 2006 of TÜV Rheinland Immissionsschutz und Energiesysteme GmbH and Addendum 936/21221709/C of 28 September 2013 of TÜV Rheinland Energie und Umwelt GmbH
- suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- the on-going surveillance of the product and the manufacturing process

**AMS designation:**

O342M

**Manufacturer:**

Environnement S.A., Poissy Cedex, France and Ansyco GmbH Karlsruhe, Germany

**Field of application:**

For continuous monitoring of ozone in ambient air.

**Measuring ranges during the performance test:**

O<sub>3</sub>    0 - 360 µg/m<sup>3</sup>  
          0 - 500 µg/m<sup>3</sup>

**Software version:**

V1.28

**Test report:**

TÜV Rheinland Immissionsschutz und Energiesysteme GmbH Cologne,  
TÜV Rheinland Group  
Report No.: 936/21205818/B vom 8 December 2006

**Notification of announcement by the German Federal Environment Agency dated 12th April 2007 (BAnz. p. 4139, Chapter III Number 3.1)**

The measuring system O342M for O<sub>3</sub>, manufactured by Environnement, fulfils the requirements of Standard EN 14625 (December 2012). Furthermore, the manufacturing proves and the quality management system of the measuring system O342M for O<sub>3</sub> fulfil the requirements of EN 15267.

The test report of the performance test with report number 936/21205818/B as well as an addendum as an integral part of to the test report with report number 936/21221709/C can be viewed on the internet at [www.qal1.de](http://www.qal1.de).

Statement by TÜV Rheinland Energie und Umwelt GmbH dated 28th September 2013

**Certified product**

This certificate applies to automated measurement systems conforming to the following description:

The analyser O342M measures ozone (O<sub>3</sub>) in ambient air. It works by absorbing ozone in the UV range. The broad UV absorption spectrum of ozone has a maximum of approx. 254 nm. This wavelength precisely coincides with the strongest emission line of a mercury low-pressure lamp of 253.7 nm as well as with the maximum spectral sensitivity of the UV detector, a caesium telluride vacuum UV diode. This is an ideal requirement for photometric measurement of ozone.

The measuring principle complies with the standard reference method as stipulated in EN 14625.

**General notes**

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied to the product or used in publicity material for the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and the validity is also accessible on the internet: [qal1.de](http://qal1.de).

Certification of Model O342M for O<sub>3</sub> is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

**Initial test:**

Test report: 936/21205818/B dated 08 December 2006  
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Cologne  
Publication: BAnz. 20 April 2007, No. 75, p. 4139, chapter III, No. 3.1  
Announcement by UBA from 12 April 2007

**Initial certification according to EN 15267:**

Certificate No. 0000040331: 29 April 2014  
Expiration date of the certificate: 31 March 2019

Test report: 936/21205818/B dated 08 December 2006  
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Cologne  
Addendum 936/21221709/C of 28 September 2013  
TÜV Rheinland Energie und Umwelt GmbH, Cologne  
Publication: BAnz AT 01 April 2014 B12, chapter VI, notification 21  
Announcement by UBA from 27 February 2014

**Notification:**

Publication: BAnz AT 01 April 2014 B12, chapter VI, notification 21  
Announcement by UBA from 27 February 2014

**Renewal of the certificate according to EN 15267:**

Certificate No. 0000040331\_01: 1 April 2019  
Expiration date of the certificate: 30 June 2020

Calculation of overall uncertainty lab test (Device 1)

Measuring device: Environment 0342M		Serial-No.:	Gerät 1	nmol/mol	
Measured component: O <sub>3</sub>		1h-alert threshold: 120			
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.400	u <sub>r,z</sub>	0.0025
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	1.000	u <sub>r,jh</sub>	0.0161
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	-0.200	u <sub>l,jh</sub>	0.0192
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	-0.010	u <sub>gp</sub>	0.0089
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.023	u <sub>gt</sub>	0.0705
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.035	u <sub>st</sub>	0.0768
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	-0.040	u <sub>v</sub>	0.2904
8a	Interferent H <sub>2</sub> O with 21 mmol/mol	≤ 10 nmol/mol (Zero) ≤ 10 nmol/mol (Span)	1.000 1.100	u <sub>H2O</sub>	0.6737
8b	Interferent Toluene with 0,5 µmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	0.100 2.600	u <sub>int,pos</sub> or	8.6700
8c	Interferent Xylene with 0,5 µmol/mol	≤ 5.0 nmol/mol (Zero) ≤ 5.0 nmol/mol (Span)	0.200 2.500	u <sub>int,neg</sub>	
9	Averaging effect	≤ 7.0% of measured value	3.000	u <sub>av</sub>	4.3200
18	Difference sample/calibration port	≤ 1.0%	0.000	u <sub>asc</sub>	0.0000
21	Uncertainty of test gas	≤ 3.0%	2.000	u <sub>cg</sub>	1.4400
Combined standard uncertainty				u <sub>c</sub>	3.9482
Expanded uncertainty				U	7.8964
Relative expanded uncertainty				W	6.58
Maximum allowed expanded uncertainty				W <sub>req</sub>	15

Calculation of overall uncertainty lab test

(Device 2)

Measuring device:		Serial-No.:		Gerät 2	
Measured component:		1h-alert threshold:		120	
Environment O342M		O <sub>3</sub>		nmol/mol	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.400	u <sub>r,z</sub>	0.0026
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	0.800	u <sub>r,lv</sub>	0.0108
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	-1.300	u <sub>l,v</sub>	0.8112
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	0.010	u <sub>gp</sub>	0.0089
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.040	u <sub>gt</sub>	0.2133
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.065	u <sub>st</sub>	0.2649
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	-0.045	u <sub>v</sub>	0.3675
8a	Interferent H <sub>2</sub> O with 21 mmol/mol	≤ 10 nmol/mol (Zero) ≤ 10 nmol/mol (Span)	0.000 0.500	u <sub>H2O</sub>	0.1392
8b	Interferent Toluene with 0.5 µmol/mol	≤ 5.0 nmol/mol (Zero)	0.100	u <sub>int,pos</sub> or	4.0833
8c	Interferent Xylene with 0.5 µmol/mol	≤ 5.0 nmol/mol (Zero)	1.500 0.400	u <sub>int,neg</sub>	
9	Averaging effect	≤ 7.0% of measured value	2.000	u <sub>av</sub>	0.0432
18	Difference sample/calibration port	≤ 1.0%	0.000	u <sub>bsc</sub>	0.0000
21	Uncertainty of test gas	≤ 3.0%	2.000	u <sub>cg</sub>	1.4400
		Combined standard uncertainty		u <sub>c</sub>	2.7175
		Expanded uncertainty		U	5.4351
		Relative expanded uncertainty		W	4.53
		Maximum allowed expanded uncertainty		W <sub>req</sub>	15

Calculation of overall uncertainty lab and field test (Device 1)

Measuring device:		Serial-No.:		Gerät 1	
Environment C0342M		1h-alert threshold:		120	
O <sub>3</sub>		Partial uncertainty		Square of partial uncertainty	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.400	u <sub>r,z</sub>	0.0025
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	1.000	u <sub>r,th</sub>	-
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	-0.200	u <sub>l,th</sub>	0.0192
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	-0.010	u <sub>sp</sub>	0.0089
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.023	u <sub>gt</sub>	0.0705
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.035	u <sub>st</sub>	0.0768
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	-0.040	u <sub>v</sub>	0.2904
8a	Interferent H <sub>2</sub> O with 21 mmol/mol	≤ 10 nmol/mol (Zero)	1.000	u <sub>H2O</sub>	0.6737
		≤ 10 nmol/mol (Span)	1.100		
8b	Interferent Toluene with 0.5 µmol/mol	≤ 5.0 nmol/mol (Zero)	0.100	u <sub>int,pos</sub>	
		≤ 5.0 nmol/mol (Span)	2.600		
8c	Interferent Xylene with 0.5 µmol/mol	≤ 5.0 nmol/mol (Zero)	0.200	or	8.6700
		≤ 5.0 nmol/mol (Span)	2.500	u <sub>int,neg</sub>	
9	Averaging effect	≤ 7.0% of measured value	3.000	u <sub>av</sub>	4.3200
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	3.240	u <sub>r</sub>	15.1165
11	Long term drift at zero level	≤ 5.0 nmol/mol	-0.580	u <sub>l,z</sub>	0.1121
12	Long term drift at span level	≤ 5.0% of max. of certification range	2.280	u <sub>l,th</sub>	2.4952
18	Difference sample/calibration port	≤ 1.0%	0.000	u <sub>Asc</sub>	0.0000
21	Uncertainty of test gas	≤ 3.0%	2.000	u <sub>cg</sub>	1.4400
Combined standard uncertainty				u <sub>c</sub>	5.7703
Expanded uncertainty				U	11.5405
Relative expanded uncertainty				W	9.62
Maximum allowed expanded uncertainty				W <sub>req</sub>	15

ty lab and field test (Device 2)

Calculation of overall uncertain-



Measuring device: Environnement O342M		Serial-No.:		Gerät 2	nmol/mol
Measured component: O <sub>3</sub>		1h-alert threshold:		120	
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.400	u <sub>r,z</sub>	0.0026
2	Repeatability standard deviation at 1h-alert threshold	≤ 3.0 nmol/mol	0.800	u <sub>r,h</sub>	-
3	"lack of fit" at 1h-alert threshold	≤ 4.0% of measured value	-1.300	u <sub>l,h</sub>	0.8112
4	Sensitivity coefficient of sample gas pressure at 1h-alert threshold	≤ 2.0 nmol/mol/kPa	0.010	u <sub>sp</sub>	0.0089
5	Sensitivity coefficient of sample gas temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.040	u <sub>gt</sub>	0.2133
6	Sensitivity coefficient of surrounding temperature at 1h-alert threshold	≤ 1.0 nmol/mol/K	0.065	u <sub>st</sub>	0.2649
7	Sensitivity coefficient of electrical voltage at 1h-alert threshold	≤ 0.30 nmol/mol/V	-0.045	u <sub>v</sub>	0.3675
8a	Interferent H <sub>2</sub> O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	0.000	u <sub>H2O</sub>	0.1392
8b	Interferent Toluene with 0.5 µmol/mol	≤ 10 nmol/mol (Span)	0.500	u <sub>int,pos</sub> or u <sub>int,neg</sub>	4.0833
		≤ 5.0 nmol/mol (Zero)	0.100		
		≤ 5.0 nmol/mol (Span)	1.500		
8c	Interferent Xylene with 0.5 µmol/mol	≤ 5.0 nmol/mol (Zero)	0.400		
		≤ 5.0 nmol/mol (Span)	2.000		
9	Averaging effect	≤ 7.0% of measured value	0.300	u <sub>av</sub>	0.0432
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	3.240	u <sub>r,f</sub>	15.1165
11	Long term drift at zero level	≤ 5.0 nmol/mol	0.790	u <sub>d,z</sub>	0.2080
12	Long term drift at span level	≤ 5.0% of max. of certification range	3.550	u <sub>d,h</sub>	6.0492
18	Difference sample/calibration port	≤ 1.0%	0.000	u <sub>asc</sub>	0.0000
21	Uncertainty of test gas	≤ 3.0%	2.000	u <sub>cg</sub>	1.4400
Combined standard uncertainty				u <sub>c</sub>	5.3617
Expanded uncertainty				U	10.7234
Relative expanded uncertainty				W	8.94
Maximum allowed expanded uncertainty				W <sub>req</sub>	15